

§ 606.40

or blood components shall be adequate in number, educational background, training and experience, including professional training as necessary, or combination thereof, to assure competent performance of their assigned functions, and to ensure that the final product has the safety, purity, potency, identity and effectiveness it purports or is represented to possess. All personnel shall have capabilities commensurate with their assigned functions, a thorough understanding of the procedures or control operations they perform, the necessary training or experience, and adequate information concerning the application of pertinent provisions of this part to their respective functions.

(c) Persons whose presence can adversely affect the safety and purity of the products shall be excluded from areas where the collection, processing, compatibility testing, storage or distribution of blood or blood components is conducted.

[40 FR 53532, Nov. 18, 1975, as amended at 49 FR 23833, June 8, 1984; 55 FR 11014, Mar. 26, 1990; 62 FR 53538, Oct. 15, 1997]

Subpart C—Plant and Facilities

§ 606.40 Facilities.

Facilities shall be maintained in a clean and orderly manner, and shall be of suitable size, construction and location to facilitate adequate cleaning, maintenance and proper operations. The facilities shall:

(a) Provide adequate space for the following when applicable:

(1) Private and accurate examinations of individuals to determine their suitability as blood donors.

(2) The withdrawal of blood from donors with minimal risk of contamination, or exposure to activities and equipment unrelated to blood collection.

(3) The storage of blood or blood components pending completion of tests.

(4) The quarantine storage of blood or blood components in a designated location pending repetition of those tests that initially gave questionable serological results.

(5) The storage of finished products prior to distribution.

21 CFR Ch. I (4–1–10 Edition)

(6) The quarantine storage, handling and disposition of products and reagents not suitable for use.

(7) The orderly collection, processing, compatibility testing, storage and distribution of blood and blood components to prevent contamination.

(8) The adequate and proper performance of all steps in plasmapheresis, plateletpheresis and leukapheresis procedures.

(9) The orderly conduction of all packaging, labeling and other finishing operations.

(b) Provide adequate lighting, ventilation and screening of open windows and doors.

(c) Provide adequate, clean, and convenient handwashing facilities for personnel, and adequate, clean, and convenient toilet facilities for donors and personnel. Drains shall be of adequate size and, where connected directly to a sewer, shall be equipped with traps to prevent back-siphonage.

(d) Provide for safe and sanitary disposal for the following:

(1) Trash and items used during the collection, processing and compatibility testing of blood and blood components.

(2) Blood and blood components not suitable for use or distribution.

Subpart D—Equipment

§ 606.60 Equipment.

(a) Equipment used in the collection, processing, compatibility testing, storage and distribution of blood and blood components shall be maintained in a clean and orderly manner and located so as to facilitate cleaning and maintenance. The equipment shall be observed, standardized and calibrated on a regularly scheduled basis as prescribed in the Standard Operating Procedures Manual and shall perform in the manner for which it was designed so as to assure compliance with the official requirements prescribed in this chapter for blood and blood products.

(b) Equipment that shall be observed, standardized and calibrated with at least the following frequency, include but are not limited to:

Equipment	Performance check	Frequency	Frequency of calibration
Temperature recorder	Compare against thermometer	Daily	As necessary.
Refrigerated centrifuge ..	Observe speed and temperature	Each day of use	Do.
Hematocrit centrifuge	Standardize before initial use, after repairs or adjustments, and annually. Timer every 3 mo.
General lab centrifuge	Tachometer every 6 mo.
Automated blood-typing machine.	Observe controls for correct results	Each day of use.	
Hemoglobinometer	Standardize against cyanmethemoglobin standard.do.	
Refractometer	Standardize against distilled waterdo.	
Blood container scale	Standardize against container of known weight.do	As necessary.
Water bath	Observe temperaturedo	Do.
Rh view boxdodo	Do.
Autoclavedo	Each time of use	Do.
Serologic rotators	Observe controls for correct results	Each day of use	Speed as necessary.
Laboratory thermometers.	Before initial use.
Electronic thermometers	Monthly.
Vacuum blood agitator ..	Observe weight of the first container of blood filled for correct results.	Each day of use	Standardize with container of known mass or volume before initial use, and after repairs or adjustments.

(c) Equipment employed in the sterilization of materials used in blood collection or for disposition of contaminated products shall be designed, maintained and utilized to ensure the destruction of contaminating microorganisms. The effectiveness of the sterilization procedure shall be no less than that achieved by an attained temperature of 121.5 °C (251 °F) maintained for 20 minutes by saturated steam or by an attained temperature of 170 °C (338 °F) maintained for 2 hours with dry heat.

[40 FR 53532, Nov. 18, 1975; 40 FR 55849, Dec. 2, 1975, as amended at 45 FR 9261, Feb. 12, 1980; 57 FR 11263, Apr. 2, 1992; 57 FR 12862, Apr. 13, 1992]

§ 606.65 Supplies and reagents.

All supplies and reagents used in the collection, processing, compatibility testing, storage and distribution of blood and blood components shall be stored in a safe, sanitary and orderly manner.

(a) All surfaces coming in contact with blood and blood components intended for transfusion shall be sterile, pyrogen-free, and shall not interact with the product in such a manner as to have an adverse effect upon the safety, purity, potency or effectiveness of the product. All final containers and closures for blood and blood components not intended for transfusion shall be clean and free of surface solids and other contaminants.

(b) Each blood collecting container and its satellite container(s), if any, shall be examined visually for damage or evidence of contamination prior to its use and immediately after filling. Such examination shall include inspection for breakage of seals, when indicated, and abnormal discoloration. Where any defect is observed, the container shall not be used, or, if detected after filling, shall be properly discarded.

(c) Representative samples of each lot of the following reagents or solutions shall be tested on a regularly scheduled basis by methods described in the Standard Operating Procedures Manual to determine their capacity to perform as required:

Reagent or solution	Frequency of testing
Anti-human globulin	Each day of use.
Blood grouping reagents	Do.
Lectins	Do.
Antibody screening and reverse grouping cells.	Do.
Hepatitis test reagents	Each run.
Syphilis serology reagents	Do.
Enzymes	Each day of use.

(d) Supplies and reagents that do not bear an expiration date shall be stored in such a manner that the oldest is used first.

(e) Supplies and reagents shall be used in a manner consistent with instructions provided by the manufacturer.