§ 90.415

§ 90.415 Raw gaseous sampling procedures.

Fit all heated sampling lines with a heated filter to extract solid particles from the flow of gas required for analysis. The sample line for HC measurement must be heated. The sample line for CO, CO_2 and NO_X analysis may be heated or unheated.

§ 90.416 Intake air flow measurement specifications.

(a) If used, the engine intake air flow measurement method used must have a range large enough to accurately measure the air flow over the engine operating range during the test. Overall measurement accuracy must be two percent of full-scale value of the measurement device for all modes except the idle mode. For the idle mode, the measurement accuracy must be ±five percent or less of the full-scale value. The Administrator must be advised of the method used prior to testing.

(b) When an engine system incorporates devices that affect the air flow measurement (such as air bleeds, air injection, pulsed air, and so forth) resulting in understated exhaust emission results, make corrections to the exhaust emission results to account for such effects.

§ 90.417 Fuel flow measurement specifications.

(a) Fuel flow measurement is required only for raw testing. Fuel flow is allowed for dilute testing.

(b) The fuel flow measurement instrument must have a minimum accuracy of one percent of full-scale flow

rate for each measurement range used. An exception is allowed for the idle mode. For this mode, the minimum accuracy is ±five percent of full-scale flow rate for the measurement range used. The controlling parameters are the elapsed time measurement of the event and the weight or volume measurement. You may apply the accuracy specifications of 40 CFR part 1065, subpart C, instead of those in this paragraph (b).

[70 FR 40449, July 13, 2005]

§ 90.418 Data evaluation for gaseous emissions.

For the evaluation of the gaseous emissions recording, record the last two minutes of each mode and determine the average values for HC, CO, CO_2 and NO_X during each mode from the average concentration readings determined from the corresponding calibration data. Longer averaging times are acceptable, but the reported sampling period must be a continuous set of data.

[70 FR 40449, July 13, 2005]

§ 90.419 Raw emission sampling calculations—gasoline fueled engines.

- (a) Derive the final weighted brakespecific mass emission rates (g/kW-hr) through the steps described in this section.
- (b) Air and fuel flow method. If both air and fuel flow mass rates are measured, use the following equations to determine the weighted emission values for the test engine:

$$\begin{split} W_{\text{NO}_{\text{X}}} &= \left(G_{\text{AIRD}} + G_{\text{FUEL}}\right) \times \frac{M_{\text{NO}_{2}}}{M_{\text{exh}}} \times \text{WNO}_{\text{X}} \times K_{\text{H}} \times \frac{1}{10^{6}} \\ W_{\text{HC}} &= \left(G_{\text{AIRD}} + G_{\text{FUEL}}\right) \times \frac{M_{\text{HC}_{\text{exh}}}}{M_{\text{exh}}} \times \text{WHC} \times \frac{1}{10^{6}} \\ W_{\text{CO}} &= \left(G_{\text{AIRD}} + G_{\text{FUEL}}\right) \times \frac{M_{\text{CO}}}{M_{\text{exh}}} \times \text{WCO} \times \frac{1}{10^{2}} \end{split}$$

Where:

W_{HC} = Mass rate of HC in exhaust [g/hr],