

Specification	PM ₁₀	PM _{2.5}			PM _{10-2.5}	
		Class I	Class II	Class III	Class II	Class III
Minimum number of candidate method samplers or analyzers per site.	3	3	3 ¹	3 ¹	3 ¹	3 ¹
Number of reference method samplers per site.	3	3	3 ¹	3 ¹	3 ¹	3 ¹
Minimum number of acceptable sample sets per site for PM ₁₀ methods:						
R _i < 60 µg/m ³	3					
R _i > 60 µg/m ³	3					
Total	10					
Minimum number of acceptable sample sets per site for PM _{2.5} and PM _{10-2.5} candidate equivalent methods:						
R _i < 30 µg/m ³ for 24-hr or R _i < 20 µg/m ³ for 48-hr samples.	3				
R _i > 30 µg/m ³ for 24-hr or R _i > 20 µg/m ³ for 48-hr samples.	3					
Each season	10	23	23	23	23	
Total, each site	10	23	23 (46 for two-season sites)	23	23 (46 for two-season sites)	
Precision of replicate reference method measurements, P _{Rj} or RP _{Rj} , respectively; RP for Class II or III PM _{2.5} or PM _{10-2.5} , maximum.	5 µg/m ³ or 7%.	2 µg/m ³ or 5%.	10% ²	10% ²	10% ²	10% ²
Precision of PM _{2.5} or PM _{10-2.5} candidate method, CP, each site.	10% ² ...	15% ² ...	15% ²	15% ²		
Slope of regression relationship.	1 ± 0.10	1 ± 0.05	1 ± 0.10	1 ± 0.10	1 ± 0.10	1 ± 0.12
Intercept of regression relationship, µg/m ³ .	0 ± 5	0 ± 1	Between: 13.55 – (15.05 × slope), but not less than –1.5; and 16.56 – (15.05 × slope), but not more than +1.5	Between: 15.05 – (17.32 × slope), but not less than –2.0; and 15.05 – (13.20 × slope), but not more than +2.0	Between: 62.05 – (70.5 × slope), but not less than –3.5; and 78.95 – (70.5 × slope), but not more than +3.5	Between: 70.50 – (82.93 × slope), but not less than –7.0; and 70.50 – (61.16 × slope), but not more than +7.0
Correlation of reference method and candidate method measurements.	≥ 0.97 ..	≥ 0.97 ..	≥ 0.93—for CCV ≤ 0.4; ≥ 0.85 + 0.2 × CCV—for 0.4 ≤ CCV ≤ 0.5; ≥ 0.95—for CCV ≥ 0.5			

¹ Some missing daily measurement values may be permitted; see test procedure.
² Calculated as the root mean square over all measurement sets.

[72 FR 32203, June 12, 2007]

TABLE C-5 TO SUBPART C OF PART 53—SUMMARY OF COMPARABILITY FIELD TESTING CAMPAIGN SITE AND SEASONAL REQUIREMENTS FOR CLASS II AND III FEMS FOR PM_{10-2.5} AND PM_{2.5}

Candidate method	Test site	A	B	C	D
PM _{2.5}	Test site location area.	Los Angeles basin or California Central Valley.	Western city such as Denver, Salt Lake City, or Albuquerque.	Midwestern city	Northeastern or mid-Atlantic city.

Candidate method	Test site	A	B	C	D
PM _{10-2.5}	Test site characteristics.	Relatively high PM _{2.5} , nitrates, and semi-volatile organic pollutants.	Cold weather, higher elevation, winds, and dust.	Substantial temperature variation, high nitrates, wintertime conditions.	High sulfate and high relative humidity.
	Class III Field test campaigns (Total: 5).	Winter and summer.	Winter only	Winter only	Summer only.
	Class II Field test campaigns (Total: 2).	Site A or B, any season		Site C or D, any season.	
	Test site location area.	Los Angeles basin or California Central Valley.	Western city such as Las Vegas or Phoenix.	Midwestern city	Large city east of the Mississippi River.
	Test site characteristics.	Relatively high PM _{2.5} , nitrates, and semi-volatile organic pollutants.	High PM _{10-2.5} to PM _{2.5} ratio, wind-blown dust.	Substantial temperature variation, high nitrates, wintertime conditions.	High sulfate and high relative humidity.
	Class III Field test campaigns (Total: 5).	Winter and summer.	Winter only	Winter only	Summer only.
Class II Field test campaigns (Total: 2).	Site A or B, any season		Site C or D, any season.		

FIGURE C-1 TO SUBPART C OF PART 53—SUGGESTED FORMAT FOR REPORTING TEST RESULTS FOR METHODS FOR SO₂, CO, O₃, NO₂

Candidate Method _____
 Reference Method _____
 Applicant _____

First Set Second Set Type 1 Hour 24 Hour

Concentration range	Date	Time	Concentration, ppm		Difference	Table C-1 spec.	Pass or fail
			Candidate	Reference			
Low ____ ppm to ____ ppm	1						
	2						
	3						
	4						
	5						
	6						
Medium ____ ppm to ____ ppm	1						
	2						
	3						