

§ 493.1299

42 CFR Ch. IV (10–1–10 Edition)

of test methods employed by the laboratory and, as applicable, the performance specifications established or verified as specified in § 493.1253. In addition, information that may affect the interpretation of test results, for example test interferences, must be provided upon request. Pertinent updates on testing information must be provided to clients whenever changes occur that affect the test results or interpretation of test results.

(f) Test results must be released only to authorized persons and, if applicable, the individual responsible for using the test results and the laboratory that initially requested the test.

(g) The laboratory must immediately alert the individual or entity requesting the test and, if applicable, the individual responsible for using the test results when any test result indicates an imminently life-threatening condition, or panic or alert values.

(h) When the laboratory cannot report patient test results within its established time frames, the laboratory must determine, based on the urgency of the patient test(s) requested, the need to notify the appropriate individual(s) of the delayed testing.

(i) If a laboratory refers patient specimens for testing—

(1) The referring laboratory must not revise results or information directly related to the interpretation of results provided by the testing laboratory;

(2) The referring laboratory may permit each testing laboratory to send the test result directly to the authorized person who initially requested the test. The referring laboratory must retain or be able to produce an exact duplicate of each testing laboratory's report; and

(3) The authorized person who orders a test must be notified by the referring laboratory of the name and address of each laboratory location where the test was performed.

(j) All test reports or records of the information on the test reports must be maintained by the laboratory in a manner that permits ready identification and timely accessibility.

(k) When errors in the reported patient test results are detected, the laboratory must do the following:

(1) Promptly notify the authorized person ordering the test and, if applicable, the individual using the test results of reporting errors.

(2) Issue corrected reports promptly to the authorized person ordering the test and, if applicable, the individual using the test results.

(3) Maintain duplicates of the original report, as well as the corrected report.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

§ 493.1299 Standard: Postanalytic systems quality assessment.

(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess and, when indicated, correct problems identified in the postanalytic systems specified in § 493.1291.

(b) The postanalytic systems quality assessment must include a review of the effectiveness of corrective actions taken to resolve problems, revision of policies and procedures necessary to prevent recurrence of problems, and discussion of postanalytic systems quality assessment reviews with appropriate staff.

(c) The laboratory must document all postanalytic systems quality assessment activities.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

Subpart L [Reserved]

Subpart M—Personnel for Nonwaived Testing

SOURCE: 57 FR 7172, Feb. 28, 1992, unless otherwise noted.

§ 493.1351 General.

This subpart consists of the personnel requirements that must be met by laboratories performing moderate complexity testing, PPM procedures, high complexity testing, or any combination of these tests.

[60 FR 20049, Apr. 24, 1995]