

TABLE 2 TO PARAGRAPH (aa)—STANDARDIZING VALUES FOR CHEMICAL RESIDUES

Class of residues	Standardizing value <sup>3</sup>
Chlorinated Hydrocarbons: <sup>1</sup>	
Aldrin .....	0.20
Benzene Hexachloride .....	0.20
Chlordane .....	0.20
Dieldrin .....	0.20
DDT .....	0.20
DDE .....	0.20
TDE .....	0.20
Endrin .....	0.20
Heptachlor .....	0.20
Heptachlor Epoxide .....	0.20
Lindane .....	0.20
Methoxychlor .....	0.20
Toxaphene .....	0.20
Hexachlorobenzene .....	0.20
Mirex .....	0.20
Nonachlor .....	0.20
Polychlorinated Biphenyls:	
Arsenic <sup>2</sup> .....	0.25
Sulfonamides <sup>2</sup> .....	0.25
Volatile Nitrosamine <sup>2</sup> .....	0.25

<sup>1</sup> Laboratory statistics are computed over all results (excluding PCB results), and for specific chemical residues.

<sup>2</sup> Laboratory statistics are only computed for specific chemical residues.

<sup>3</sup> The standardizing value of all initial accreditation and probationary check samples computations is 0.15.

(bb) *Suspension of accreditation*—Action taken by FSIS against a laboratory that temporarily removes the laboratory's right to analyze official samples. Suspension of accreditation ends when accreditation either is fully restored or is revoked.

(cc) *Systematic laboratory difference*—A comparison of one laboratory's results with the comparison mean for samples that show, on average, a consistent relationship. A laboratory that is reporting, on average, numerically greater results than the comparison mean has a positive systematic laboratory difference. Conversely, numerically smaller results indicate a negative systematic laboratory difference.

(dd) *Variability*—Random fluctuations in a laboratory's processes that cause its analytical results to deviate from a true value.

(ee) *Variance*—The expected average of the squared differences of sample results from an expected sample mean.

**§ 439.5 Applications for accreditation.**

(a) Application for accreditation shall be made on designated paper or electronic forms provided by FSIS, or otherwise in writing, by the owner or manager of a non-Federal analytical laboratory. The forms shall be sent to

the ALP or may be submitted electronically when so provided for by FSIS. The application shall specify the kinds of accreditation that are wanted by the owner or manager of the laboratory. A laboratory whose accreditation has been refused or revoked may re-apply for accreditation after 60 days from the effective date of that action, and must provide written documentation specifying what corrections were made.

(b) At the time that an Application for Accreditation is filed with the ALP, the management of a laboratory shall, for each accreditation sought, submit a check, bank draft, or money order in the amount specified in 9 CFR 391.5, made payable to the U.S. Department of Agriculture, along with the completed application for the accreditation(s). When so provided for by FSIS, electronic transfer of funds may be accepted.

(c) Accreditation will not be granted or continued, without further procedure, for failure to pay the accreditation fee(s). The fee(s) paid will be non-refundable and will be credited to the account from which the expenses of the laboratory accreditation program are paid.

(d) Annually on the anniversary date of each accreditation, FSIS will issue a bill in the amount specified in 9 CFR 391.5 for each accreditation held. Bills are payable upon receipt by check, bank draft, or money order made payable to the U.S. Department of Agriculture and become delinquent 30 days from the date of the bill.

(e) Accreditation will be terminated without further procedure for having a delinquent account. The fee(s) paid will be nonrefundable and will be credited to the account from which the expenses of the ALP are paid.

**§ 439.10 Criteria for obtaining accreditation.**

(a) Analytical laboratories may be accredited for the analyses of food chemistry analytes, as defined in § 439.1 of this part, or a specific chemical residue or a class of chemical residues in raw or processed meat and poultry products.

(b) Accreditation will be given only if the applying laboratory successfully