

### § 442.3

3.18 Bundles and Bags of Firewood

#### CHAPTER 4—TEST PROCEDURES—PACKAGES LABELED BY COUNT, LINEAR MEASURE, AREA, THICKNESS, AND COMBINATIONS OF QUANTITIES

4.5 Paper Plates and Sanitary Paper Products

4.6 Special Test Requirements for Packages Labeled by Linear or Square Measure (Area)

4.7 Polyethylene sheeting

4.8 Packages Labeled by Linear or Square (Area) Measure

4.9 Bailer Twine—Test Procedure for Length

4.10 Procedure for Checking the Area Measurement of Chamois Appendix C Glossary—wet tare

#### § 442.3 Scale requirements for accurate weights, repairs, adjustments, and replacements after inspection.

(a) All scales used to determine the net weight of meat and poultry products sold or otherwise distributed in commerce in federally inspected meat and poultry establishments will be installed, maintained, and operated in a manner that ensures accurate weights. Such scales shall meet the applicable requirements contained in National Institute of Standards and Technology (NIST) Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices," 1999 Edition, November 1988, which is incorporated by reference. This incorporation was approved by the Director of the Federal Register in accordance with 5 U.S.C. 552(a) and 1 CFR part 51. (These materials are incorporated as they exist on the date of approval.) A notice of any change in the Handbook cited here will be published in the FEDERAL REGISTER. Copies may be purchased from the Superintendent of Documents, Government Printing Office, Washington, DC 20402. The incorporation information also is available for inspection at the Office of the Federal Register Information Center, 800 North Capitol Street, NW., suite 700, Washington, DC 20408.

(b) All scales used to determine the net weight of meat or poultry products sold or otherwise distributed in commerce or in States designated under section 301(c) of the Federal Meat Inspection Act and section 5(c) of the Poultry Products Inspection Act shall

### 9 CFR Ch. III (1–1–12 Edition)

be of sufficient capacity to weigh the entire unit or package.

(c) No scale will be used at a federally inspected establishment to determine the net weight of meat or poultry products unless it has been found upon test and inspection, as specified in NIST Handbook 44 to provide accurate weight. If a scale is inspected or tested and found to be inaccurate, or if any repairs, adjustments, or replacements are made to a scale, it shall not be used until it has been reinspected and retested by a USDA official, or a State or local government weights and measures official, or a State registered or licensed scale repair firm or person, and it must meet all accuracy requirements as specified in NIST Handbook 44. If a USDA inspector has put a "Retain" tag on a scale, the tag can only be removed by a USDA inspector. As long as the tag is on the scale, it shall not be used.

#### § 442.4 Testing of scales.

(a) The operator of each official establishment that weighs meat or poultry food products will cause such scales to be tested for accuracy in accordance with the technical requirements of NIST Handbook 44, at least once during the calendar year. In cases where the scales are found not to maintain accuracy between tests, more frequent tests may be required and verified by an authorized USDA program official.

(b) The operator of each official establishment shall display on or near each scale a valid certification of the scale's accuracy from a State or local government's weights and measures authority or from a State registered or licensed scale repair firm or person, or shall have alternative documented procedures showing that the scale has been tested for accuracy in accordance with the requirements of NIST Handbook 44.

#### § 442.5 Handling of failed product.

Any lot of product that is found to be out of compliance with net weight requirements upon testing in accordance with the methods prescribed in § 442.2 of this subchapter shall be handled as follows:

(a) A lot tested in an official establishment and found not to comply with