

(b) Records of patient testing including, if applicable, instrument print-outs, must be retained.

§ 493.1289 Standard: Analytic systems quality assessment.

(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and when indicated, correct problems identified in the analytic systems specified in §§ 493.1251 through 493.1283.

(b) The analytic systems quality assessment must include a review of the effectiveness of corrective actions taken to resolve problems, revision of policies and procedures necessary to prevent recurrence of problems, and discussion of analytic systems quality assessment reviews with appropriate staff.

(c) The laboratory must document all analytic systems quality assessment activities.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

POSTANALYTIC SYSTEMS

§ 493.1290 Condition: Postanalytic systems.

Each laboratory that performs non-waived testing must meet the applicable postanalytic systems requirements in § 493.1291 unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7) that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the postanalytic systems and correct identified problems as specified in § 493.1299 for each specialty and subspecialty of testing performed.

§ 493.1291 Standard: Test report.

(a) The laboratory must have an adequate manual or electronic system(s) in place to ensure test results and other patient-specific data are accurately and reliably sent from the point of data entry (whether interfaced or entered manually) to final report destination, in a timely manner. This includes the following:

(1) Results reported from calculated

(2) Results and patient-specific data electronically reported to network or interfaced systems.

(3) Manually transcribed or electronically transmitted results and patient-specific information reported directly or upon receipt from outside referral laboratories, satellite or point-of-care testing locations.

(b) Test report information maintained as part of the patient's chart or medical record must be readily available to the laboratory and to CMS or a CMS agent upon request.

(c) The test report must indicate the following:

(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number.

(2) The name and address of the laboratory location where the test was performed.

(3) The test report date.

(4) The test performed.

(5) Specimen source, when appropriate.

(6) The test result and, if applicable, the units of measurement or interpretation, or both.

(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.

(d) Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.

(e) The laboratory must, upon request, make available to clients a list of test methods employed by the laboratory and, as applicable, the performance specifications established or verified as specified in § 493.1253. In addition, information that may affect the interpretation of test results, for example test interferences, must be provided upon request. Pertinent updates on testing information must be provided to clients whenever changes occur that affect the test results or interpretation of test results.

(f) Except as provided in § 493.1291(1), test results must be released only to authorized persons and, if applicable,