

(11) Imminently life-threatening test results, or panic or alert values.

(12) Pertinent literature references.

(13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life-threatening results, or panic, or alert values.

(14) Description of the course of action to take if a test system becomes inoperable.

(c) Manufacturer's test system instructions or operator manuals may be used, when applicable, to meet the requirements of paragraphs (b)(1) through (b)(12) of this section. Any of the items under paragraphs (b)(1) through (b)(12) of this section not provided by the manufacturer must be provided by the laboratory.

(d) Procedures and changes in procedures must be approved, signed, and dated by the current laboratory director before use.

(e) The laboratory must maintain a copy of each procedure with the dates of initial use and discontinuance as described in § 493.1105(a)(2).

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

§ 493.1252 Standard: Test systems, equipment, instruments, reagents, materials, and supplies.

(a) Test systems must be selected by the laboratory. The testing must be performed following the manufacturer's instructions and in a manner that provides test results within the laboratory's stated performance specifications for each test system as determined under § 493.1253.

(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following:

(1) Water quality.

(2) Temperature.

(3) Humidity.

(4) Protection of equipment and instruments from fluctuations and inter-

ruptions in electrical current that adversely affect patient test results and test reports.

(c) Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following:

(1) Identity and when significant, titer, strength or concentration.

(2) Storage requirements.

(3) Preparation and expiration dates.

(4) Other pertinent information required for proper use.

(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.

(e) Components of reagent kits of different lot numbers must not be interchanged unless otherwise specified by the manufacturer.

§ 493.1253 Standard: Establishment and verification of performance specifications.

(a) *Applicability.* Laboratories are not required to verify or establish performance specifications for any test system used by the laboratory before April 24, 2003.

(b)(1) *Verification of performance specifications.* Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results:

(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics:

(A) Accuracy.

(B) Precision.

(C) Reportable range of test results for the test system.

(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.

(2) *Establishment of performance specifications.* Each laboratory that modifies an FDA-cleared or approved test system, or introduces a test system not subject to FDA clearance or approval (including methods developed in-house and standardized methods such as text

§ 493.1254

book procedures), or uses a test system in which performance specifications are not provided by the manufacturer must, before reporting patient test results, establish for each test system the performance specifications for the following performance characteristics, as applicable:

- (i) Accuracy.
- (ii) Precision.
- (iii) Analytical sensitivity.
- (iv) Analytical specificity to include interfering substances.
- (v) Reportable range of test results for the test system.
- (vi) Reference intervals (normal values).
- (vii) Any other performance characteristic required for test performance.

(3) *Determination of calibration and control procedures.* The laboratory must determine the test system's calibration procedures and control procedures based upon the performance specifications verified or established under paragraph (b)(1) or (b)(2) of this section.

(c) *Documentation.* The laboratory must document all activities specified in this section.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

§ 493.1254 Standard: Maintenance and function checks.

(a) *Unmodified manufacturer's equipment, instruments, or test systems.* The laboratory must perform and document the following:

(1) Maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.

(2) Function checks as defined by the manufacturer and with at least the frequency specified by the manufacturer. Function checks must be within the manufacturer's established limits before patient testing is conducted.

(b) *Equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer.* The laboratory must do the following:

(1)(i) Establish a maintenance protocol that ensures equipment, instrument, and test system performance

42 CFR Ch. IV (10-1-22 Edition)

that is necessary for accurate and reliable test results and test result reporting.

(ii) Perform and document the maintenance activities specified in paragraph (b)(1)(i) of this section.

(2)(i) Define a function check protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting.

(ii) Perform and document the function checks, including background or baseline checks, specified in paragraph (b)(2)(i) of this section. Function checks must be within the laboratory's established limits before patient testing is conducted.

§ 493.1255 Standard: Calibration and calibration verification procedures.

Calibration and calibration verification procedures are required to substantiate the continued accuracy of the test system throughout the laboratory's reportable range of test results for the test system. Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following:

(a) Perform and document calibration procedures—

(1) Following the manufacturer's test system instructions, using calibration materials provided or specified, and with at least the frequency recommended by the manufacturer;

(2) Using the criteria verified or established by the laboratory as specified in § 493.1253(b)(3)—

(i) Using calibration materials appropriate for the test system and, if possible, traceable to a reference method or reference material of known value; and

(ii) Including the number, type, and concentration of calibration materials, as well as acceptable limits for and the frequency of calibration; and

(3) Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification.

(b) Perform and document calibration verification procedures—

(1) Following the manufacturer's calibration verification instructions;