

Application accepted by Commissioner of Customs: June 6, 1996.

Docket Number: 96-064. Applicant: University of California, Davis, Department of Geology, Davis, CA 95616. Instrument: Magnetometer and Demagnetizer. Manufacturer: Molspin Instruments, United Kingdom. Intended Use: The instruments will be used for the study of the magnetic properties of sedimentary rocks retrieved as part of the Cape Roberts Drilling Project, a collaboration that will provide information about the climatic and tectonic history of the Antarctic continent during the past 65 million years. The magnetometer will be used to determine the direction of magnetization of rocks and the alternating field demagnetizer will be used to determine whether the directions measured with the magnetometer are reliable. Application accepted by Commissioner of Customs: June 9, 1996.

Docket Number: 96-065. Applicant: University of Massachusetts, Amherst, Amherst, MA 01003. Instrument: Electron Microscope, Model JEM-3010. Manufacturer: JEOL Ltd., Japan. Intended Use: The instrument will be used to investigate the morphology and structure of polymers, proteins, ceramics, electronic materials and metals. In addition, the instrument will be used for educational purposes in the courses PSE 899 Ph.D. Dissertation, PSE 699 Master's Thesis, PSE 721 Morphology of Polymers and PSE 602 Polymer Characterization Laboratory. Application accepted by Commissioner of Customs: June 11, 1996.

Frank W. Creel,
Director, Statutory Import Programs Staff.
[FR Doc. 96-16617 Filed 6-28-96; 8:45 am]
BILLING CODE 3510-DS-P

The Pennsylvania State University, et al. Notice of Consolidated Decision on Applications for Duty-Free Entry of Scientific Instruments

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 AM and 5:00 PM in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 95-122. Applicant: The Pennsylvania State University, University Park, PA 16802. Instrument: Trace Gas Preconcentrator. Manufacturer: Finnigan MAT, Germany. Intended Use: See notice at 61 FR 6629,

February 21, 1996. Advice received from: The National Institutes of Health, March 20, 1996.

Docket Number: 95-123. Applicant: Carnegie Institution of Washington, Washington, DC 20015-1305. Instrument: Upgrade of 252 Mass Spectrometer. Manufacturer: Finnigan MAT, Germany. Intended Use: See notice at 61 FR 6629, February 21, 1996. Advice received from: The National Institutes of Health, March 20, 1996.

Docket Number: 95-128. Applicant: University of Maryland at College Park, College Park, MD 20742. Instrument: Extended SpectraKinetics Photomultiplier, Model SK.1E. Manufacturer: Applied Photophysics, United Kingdom. Intended Use: See notice at 61 FR 6630, February 21, 1996. Advice received from: The National Institutes of Health, March 21, 1996.

Docket Number: 95-129. Applicant: Massachusetts Institute of Technology, Cambridge, MA 02139. Instrument: Rapid Scanning Diode Array, Model MG 6040. Manufacturer: Hi-Tech Scientific, United Kingdom. Intended Use: See notice at 61 FR 6630, February 21, 1996. Advice received from: The National Institutes of Health, March 21, 1996.

Docket Number: 95-130. Applicant: University of Wisconsin-Madison, Madison, WI 53706. Instrument: Upgraded Pulse Compressor, Model DMP-100. Manufacturer: Microlase Optical Systems Ltd., United Kingdom. Intended Use: See notice at 61 FR 6630, February 21, 1996. Advice received from: The National Institutes of Health, March 21, 1996.

Docket Number: 96-001. Applicant: University of California, Davis, Davis, CA 95616-8605. Instrument: Water Gas Phase Equilibration System. Manufacturer: Finnigan MAT, Germany. Intended Use: See notice at 61 FR 8041, March 1, 1996. Advice received from: The National Institutes of Health, March 25, 1996.

Docket Number: 96-013. Applicant: Northern Illinois University, DeKalb, IL 60115-2862. Instrument: Diode-Array Detector for Stopped-Flow Spectrometer, Model PDA.1. Manufacturer: Applied Photophysics, Ltd., United Kingdom. Intended Use: See notice at 61 FR 11614, March 21, 1996. Advice received from: The National Institutes of Health, March 27, 1996.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instruments, for the purposes for which the instruments are intended to be used, is being manufactured in the United States. Reasons: These are compatible accessories for instruments previously

imported for the use of the applicants. In each case, the instrument and accessory were made by the same manufacturer. The National Institutes of Health advises that the accessories are pertinent to the intended uses and that it knows of no comparable domestic accessories.

We know of no domestic accessories which can be readily adapted to the previously imported instruments.

Frank W. Creel,
Director, Statutory Import Programs Staff.
[FR Doc. 96-16615 Filed 6-28-96; 8:45 am]
BILLING CODE 3510-DS-P

University of California, Los Alamos; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 96-008. Applicant: University of California, Los Alamos National Laboratory, Los Alamos, CA 87545. Instrument: Mass Spectrometer, Model Plasma Trace 2. Manufacturer: Fisons Instruments, United Kingdom. Intended Use: See notice at 61 FR 8042, March 1, 1996.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides a magnet sector analyzer with sub-ppt detection limits for Be, Co, In, Bi and U with abundance sensitivities of 5×10^{-7} for Zn and 1×10^{-6} for U at resolution 400. These capabilities are pertinent to the applicant's intended purposes and we know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,
Director, Statutory Import Programs Staff.
[FR Doc. 96-16616 Filed 6-28-96; 8:45 am]
BILLING CODE 3510-DS-P