

Parties to this proceeding may request disclosure within five days of publication of this notice and any interested party may request a hearing within 10 days of publication. Any hearing, if requested, will be held 44 days after the date of publication, or the first working day thereafter. Interested parties may submit case briefs and/or written comments no later than 30 days after the date of publication. Rebuttal briefs and rebuttals to written comments, limited to issues raised in such briefs or comments, may be filed no later than 37 days after the date of publication. The Department issue the final results of the administrative review, which will include the results of its analysis of issues raised in any such written comments or at the hearing, within 120 days from the date of publication of these preliminary results.

The Department shall determine, and Customs shall assess, antidumping duties on all appropriate entries. Individual differences between USP and NV may vary from the percentages stated above. The Department will issue appraisal instructions directly to Customs. The final results of this review shall be the basis for the assessment of antidumping dumping duties on entries of merchandise covered by the determination and for future deposits of estimated duties.

Furthermore, the following deposit requirements will be effective upon completion of the final results of this administrative review for all shipments of titanium sponge from Russia entered, or withdrawn from warehouse, for consumption on or after the publication date of the final results of these administrative reviews, as provided by section 751(a)(1) of the Act: (1) The cash deposit rates for Interlink, TMC, Cometals, and AVISMA will be the rates established in the final results of this administrative review; (2) for merchandise exported by manufacturers or exporters not covered in this review but covered in the original LTFV investigation or a previous review and have a separate rate, the cash deposit rate will continue to be the most recent rate published in the final determination or final results for which the manufacturer or exporter received a company-specific rate; (3) for Russian manufacturers or exporters not covered in the LTFV investigation or in this or prior administrative reviews, the cash deposit rate will continue to be the Russia-wide rate; and (4) the cash deposit rate for non-Russian exporters of subject merchandise from Russia who were not covered in the LTFV investigation or in this or prior administrative reviews will be the rate

applicable to the Russian supplier of that exporter. These deposit rates, when imposed, shall remain in effect until publication of the final results of the next administrative review.

This notice serves as a preliminary reminder to importers of their responsibility under 19 CFR 353.26(b) to file a certificate regarding the reimbursement of antidumping duties prior to liquidation of the relevant entries during these review periods. Failure to comply with this requirement could result in the Secretary's presumption that reimbursement of antidumping duties occurred and the subsequent assessment of double antidumping duties.

This administrative review and notice are in accordance with section 751(a)(1) of the Act (19 U.S.C. 1675(a)(1)).

Dated: May 5, 1997.

Robert S. LaRussa,

Acting Assistant Secretary for Import Administration.

[FR Doc. 97-12202 Filed 5-9-97; 8:45 am]

BILLING CODE 3510-DS-P

DEPARTMENT OF COMMERCE

International Trade Administration

Norfolk State University; Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 a.m. and 5 p.m. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, NW., Washington, DC.

Docket Number: 97-013. Applicant: Norfolk State University, Norfolk, VA 23504. Instrument: Q-Band ESR Spectrometer with Accessories. Manufacturer: Bruker Instruments Inc., Germany. Intended Use: See notice at 62 FR 10543, March 7, 1997.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides: (1) A 34 GHz Gunn source (Q-Band) with 80 mW power, 60 dB +/- 0.2 attenuation, 100 MHz tuning range and (2) a low temperature Q-Band resonator. A domestic spectrometer manufacturer advises that (1) these capabilities are pertinent to the applicant's intended purpose and (2) it

knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use.

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

Director, Statutory Import Programs Staff.

[FR Doc. 97-12390 Filed 5-9-97; 8:45 am]

BILLING CODE 3510-DS-P

DEPARTMENT OF COMMERCE

International Trade Administration

Oregon Graduate Institute of Science and Technology, et al; Consolidated Decision on Applications for Duty-Free Entry of Scientific Instruments

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 a.m. and 5 p.m. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, NW., Washington, DC.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instruments described below, for such purposes as each is intended to be used, is being manufactured in the United States.

Docket Number: 96-141. Applicant: Oregon Graduate Institute of Science and Technology, Portland, OR 97291-1000. Instrument: Stopped-Flow Spectrometer, Model SX.18MV. Manufacturer: Applied Photophysics, Ltd., United Kingdom. Intended Use: See notice at 62 FR 5619, February 6, 1997. Reasons: The foreign instrument provides: (1) Multimixing capability, (2) a deadtime of 1.2 ms permitting analysis of reaction rates to 1500 s⁻¹ and (3) a photo-diode array detector. Advice received from: National Institutes of Health, March 19, 1997.

Docket Number: 97-017. Applicant: University of California, San Diego, La Jolla, CA 92093-0931. Instrument: Sleep Recorder, Model Vitaport 2. Manufacturer: TEMEC Instruments BV, The Netherlands. Intended Use: See notice at 62 FR 13600, March 21, 1997. Reasons: The foreign instrument provides: (1) Electronic measurements of electrophysical (e.g. EEG and EOG) and cardiorespiratory (e.g. ECG and RIP-THOR) parameters and (2) minimized weight power consumption and physical dimensions appropriate for