

**DEPARTMENT OF LABOR****Occupational Safety and Health Administration**

[Docket Nos. NRTL1-88, NRTL1-89, NRTL2-90, NRTL3-90, NRTL2-92, NRTL3-92, NRTL1-93, NRTL2-93, NRTL3-93, NRTL4-93, NRTL1-97, NRTL1-98, NRTL1-99, NRTL1-2001, NRTL2-2001]

**Modify Scope of Recognition of NRTLs**

**AGENCY:** Occupational Safety and Health Administration (OSHA), Labor.

**ACTION:** Notice.

**SUMMARY:** This notice modifies the scope of recognition of certain Nationally Recognized Testing Laboratories (NRTLs).

**EFFECTIVE DATE:** January 6, 2003.

**FOR FURTHER INFORMATION CONTACT:**

Bernard Pasquet, Office of Technical Programs and Coordination Activities, NRTL Program, Occupational Safety and Health Administration, U.S. Department of Labor, 200 Constitution Avenue, NW., Room N3653, Washington, DC 20210, or phone (202) 693-2110.

**SUPPLEMENTARY INFORMATION:** The Occupational Safety and Health Administration (OSHA) hereby gives notice of changes to the scope of recognition of the Nationally Recognized Testing Laboratories (NRTLs) listed below. Specifically, some of the test standards that OSHA currently includes in the scope of recognition of these NRTLs are no longer "appropriate test standards" primarily because they have been withdrawn or replaced. As a result, we will delete them from the scope of recognition of each affected NRTL, as listed below in this notice. This modification in scope will be noted by making appropriate changes to the listing of test standards in our informational Web page for each NRTL, which detail OSHA's official scope of recognition for the NRTL. These Web pages can be accessed at <http://www.osha-slc.gov/dts/otpc/nrtl/index.html>. In this notice, we list the test standards to be removed for each NRTL below under the heading "Withdrawn or Replaced Standards." We provide the following information for those who may be unfamiliar with OSHA requirements concerning NRTLs.

OSHA recognition of any NRTL signifies that the organization has met the legal requirements in section 1910.7 of Title 29, Code of Federal Regulations (29 CFR 1910.7). Recognition is an acknowledgment that the organization can perform independent safety testing and certification of the specific products covered within its scope of recognition

and is not a delegation or grant of government authority. As a result of recognition, employers may use products "properly certified" by the NRTL to meet OSHA standards that require testing and certification.

In testing and certifying (*i.e.*, approving) such products, NRTLs must demonstrate that the products conform to "appropriate test standards." This term is defined under 29 CFR 1910.7(c) and essentially means consensus-based product safety test standards developed and maintained current by U.S.-based standards developing organizations (SDOs). Such test standards are not OSHA standards, which are general requirements that employers must meet, but, individually, specify technical safety requirements that a particular type of product must meet.

OSHA recognizes each NRTL for a particular scope of recognition, which includes a list of those product safety test standards that the NRTL may use in approving products. As a normal part of its operations, an SDO occasionally withdraws existing test standards or adopts replacement test standards. In such cases, OSHA can no longer consider the withdrawn or replaced standards as "appropriate," and as a result, the Agency can no longer recognize NRTLs for the standard.

To substitute other test standards for those we are removing, under our policy, the NRTL may request or OSHA can provide recognition for comparable test standards, *i.e.*, other appropriate test standards covering comparable product testing. We list these test standards below for each NRTL under the headings "Comparable Replacement Standards." In one case (UL 1598), OSHA already includes the comparable test standard in the NRTL's scope. However, in many cases, there is no replacement standard or the NRTL did not request one prior to publication of this notice. However, if we receive such a request after publication of this notice and determine the test standard is "comparable," as described above, OSHA will add it to the NRTL's scope of recognition and therefore to OSHA's informational web page for the NRTL.

**Applied Research Laboratories, Inc. (ARL)**

[Docket No. NRTL1-97]

**Withdrawn or Replaced Standards**

UL 1570 Fluorescent Lighting Fixtures  
 UL 1571 Incandescent Lighting Fixtures  
 UL 1572 High Intensity Discharge Lighting Fixtures

**Comparable Replacement Standards (If Applicable)**

UL 1598 Luminaries (already included in NRTL's scope)

**Canadian Standards Association (CSA)**

[Docket No. NRTL2-92]

**Withdrawn or Replaced Standards**

ANSI C37.71 Three Phase, Manually Operated Subsurface Load Interrupting Switches for Alternating-Current Systems  
 ANSI Z21.10.2 Water Heaters—Sidearm Type Water Heaters  
 ANSI Z21.11.1 Gas-Fired Room Heaters—Vented Room Heaters  
 ANSI Z21.48 Gas-Fired Gravity and Fan Type Floor Furnaces  
 ANSI Z21.49 Gas-Type Gravity and Fan Type Vented Wall Furnaces  
 ANSI Z83.3 Gas Utilization Equipment in Large Boilers  
 ANSI Z83.17 Direct Gas Fired Door Heaters  
 UL 198B Class H Fuses  
 UL 198C High-Interrupting-Capacity Fuses, Current Limiting Type  
 UL 198D Class K Fuses  
 UL 198E Class R Fuses  
 UL 198F Plug Fuses  
 UL 198G Fuse for Supplementary Overcurrent Protection  
 UL 198H Class T Fuses  
 UL 198L D-C Fuses for Industrial Use  
 ANSI/NEMA 250 Enclosures for Electrical Equipment  
 UL 910 Test for Flame-Propagation and Smoke-Density Values for Electrical and Optical-Fiber Cables Used in Spaces Transporting Environmental Air  
 UL 1087 Molded-Case Switches  
 UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*  
 UL 1270 Radio Receivers, Audio Systems, and Accessories  
 UL 1409 Low-Voltage Video Products Without Cathode-Ray-Tube Displays  
 UL 1410 Television Receivers and High-Voltage Video Products  
 UL 1570 Fluorescent Lighting Fixtures  
 UL 1571 Incandescent Lighting Fixtures  
 UL 1572 High Intensity Discharge Lighting Fixtures  
 UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements  
 UL 3101-2-20 Electrical Equipment for Laboratory Use; Part 2: Laboratory Centrifuges  
 UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*  
 UL 3121-1 Process Control Equipment  
 UL 8730-1 Electrical Controls for Household and Similar Use; Part 1: General Requirements

- UL 8730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 8730-2-4 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Motor Compressors or Hermetic and Semi-Hermetic Type\*\*
- UL 8730-2-6 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Pressure Sensing Controls Including Mechanical Requirements
- UL 8730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches
- UL 8730-2-8 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Water Valves
- UL 8730-2-9 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Temperature Sensing Controls\*\*
- UL 8730-2-14 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electric Actuators
- UL 60730-2-10 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically-Operated Motor Starting Relays
- UL 60730-2-11 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Energy Regulators
- UL 60730-2-12 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically-Operated Doors
- UL 60730-2-13 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Humidity Sensing Controls
- UL 60730-2-16 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level-Operating Controls of the Float Type for Household and Similar Applications
- Comparable Replacement Standards (If Applicable)*
- UL 1598 Luminaries (already included in NRTL's scope)
- UL 60730-1A Automatic Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 60730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls
- UL 61010A-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- FM Global Technologies LLC (FMGT) [formerly Factory Mutual Research Corporation]**
- [Docket No. NRTL3-93]
- Withdrawn or Replaced Standards*
- ANSI S12.12 Nonincendive Electrical Equipment for Use in Class I and II, Division 2, and Class III, Divisions 1 and 2, Hazardous (Classified) Locations
- ANSI/NEMA 250 Enclosures for Electrical Equipment
- for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 60730-2-6 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Pressure Sensing Controls Including Mechanical Requirements
- UL 60730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches
- UL 60730-2-10A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Motor Starting Relays
- UL 60730-2-11A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Energy Regulators
- UL 60730-2-12A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Door Locks
- UL 60730-2-13A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Humidity Sensing Controls
- UL 60730-2-14 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electric Actuators
- UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls
- UL 61010A-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 61010A-2-020 Electrical Equipment for Laboratory Use; Part 2: Particular Requirements for Laboratory Centrifuges
- UL 61010C-1 Process Control Equipment
- Curtis-Straus LLC (CSL)**
- [Docket No. NRTL-1-99]
- Withdrawn or Replaced Standards*
- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- Comparable Replacement Standards (If Applicable)*
- UL 61010A-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- Entela, Inc. (ENT)**
- [Docket No. NRTL2-93]
- Withdrawn or Replaced Standards*
- UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*
- UL 1270 Radio Receivers, Audio Systems, and Accessories
- UL 1410 Television Receivers and High-Voltage Video Products
- UL 1570 Fluorescent Lighting Fixtures
- UL 1571 Incandescent Lighting Fixtures
- UL 1572 High Intensity Discharge Lighting Fixtures
- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*
- UL 8730-1 Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 8730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 8730-2-4 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Motor Compressors or Hermetic and Semi-Hermetic Type\*\*
- UL 8730-2-8 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Water Valves
- Comparable Replacement Standards (If Applicable)*
- UL 1598 Luminaries (already included in NRTL's scope)
- UL 60730-1A Automatic Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 60730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls
- UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

*Comparable Replacement Standards (If Applicable)*

ANSI 12.12.01 Nonincendive Electrical Equipment for Use in Class I and II, Division 2 and Class III, Divisions 1 and 2 Hazardous (Classified) Locations

**Intertek Testing Services NA, Inc. (ITSNA)**

[Docket No. NRTL1-89]

*Withdrawn or Replaced Standards*

ANSI S12.12 Nonincendive Electrical Equipment for Use in Class I and II, Division 2, and Class III, Divisions 1 and 2, Hazardous (Classified) Locations

ANSI Z21.11.1 Gas-Fired Room Heaters—Vented Room Heaters

ANSI Z21.48 Gas-Fired Gravity and Fan Type Floor Furnaces

ANSI Z21.49 Gas-Type Gravity and Fan Type Vented Wall Furnaces

UL 136 Pressure Cookers (not appropriate—no NRTL approval requirement)

UL 198B Class H Fuses

UL 198D Class K Fuses

UL 198E Class R Fuses

UL 198F Plug Fuses

UL 198G Fuse for Supplementary Overcurrent Protection

UL 198H Class T Fuses

UL 198L D-C Fuses for Industrial Use

ANSI/NEMA 250 Enclosures for Electrical Equipment

UL 900 Air-Filter Units (not appropriate—no NRTL approval requirement)

UL 910 Test for Flame-Propagation and Smoke-Density Values for Electrical and Optical-Fiber Cables Used in Spaces Transporting Environmental Air

UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*

UL 1270 Radio Receivers, Audio Systems, and Accessories

UL 1409 Low-Voltage Video Products Without Cathode-Ray-Tube Displays

UL 1410 Television Receivers and High-Voltage Video Products

UL 1570 Fluorescent Lighting Fixtures

UL 1571 Incandescent Lighting Fixtures

UL 1572 High Intensity Discharge Lighting Fixtures

UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements

UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

UL 8730-1 Electrical Controls for Household and Similar Use; Part 1: General Requirements

UL 8730-2-3 Automatic Electrical Controls for Household and Similar

Use; Part 2: Particular Requirements for Thermal Motor Protectors for Ballasts for Tubular Fluorescent Lamps

UL 8730-2-4 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Motor Compressors or Hermetic and Semi-Hermetic Type\*\*

UL 8730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches

UL 8730-2-8 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Water Valves

*Comparable Replacement Standards (If Applicable)*

ANSI 12.12.01 Nonincendive Electrical Equipment for Use in Class I and II, Division 2 and Class III, Divisions 1 and 2 Hazardous (Classified) Locations

UL 1598 Luminaries (already included in NRTL's scope)

UL 60730-1A Automatic Electrical Controls for Household and Similar Use; Part 1: General Requirements

UL 60730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps

UL 60730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches

UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls

UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**MET Laboratories, Inc. (MET)**

[Docket No. NRTL1-88]

*Withdrawn or Replaced Standards*

UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*

UL 1270 Radio Receivers, Audio Systems, and Accessories

UL 1409 Low-Voltage Video Products Without Cathode-Ray-Tube Displays

UL 1410 Television Receivers and High-Voltage Video Products

UL 1570 Fluorescent Lighting Fixtures

UL 1571 Incandescent Lighting Fixtures

UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements

UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

*Comparable Replacement Standards (If Applicable)*

UL 1598 Luminaries (already included in NRTL's scope)

UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**National Technical Systems, Inc. (NTS)**

[Docket No. NRTL1-98]

*Withdrawn or Replaced Standards*

UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements

UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

*Comparable Replacement Standards (If Applicable)*

UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**SGS U.S. Testing Company, Inc. (SGSUS)**

[Docket No. NRTL2-90]

*Withdrawn or Replaced Standards*

UL 1270 Radio Receivers, Audio Systems, and Accessories

UL 1571 Incandescent Lighting Fixtures

UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements

UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

*Comparable Replacement Standards (If Applicable)*

UL 1492 Audio-Video Products and Accessories

UL 1598 Luminaries (already included in NRTL's scope)

UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**Southwest Research Institute (SWRI)**

[Docket No. NRTL-3-90]

*Withdrawn or Replaced Standards*

UL 910 Test for Flame-Propagation and Smoke-Density Values for Electrical and Optical-Fiber Cables Used in Spaces Transporting Environmental Air

*Comparable Replacement Standards (If Applicable)*

None

**TUV America, Inc. (TUVAM)**

[Docket No. NRTL-2-2001]

*Withdrawn or Replaced Standards*

- UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*
- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

*Comparable Replacement Standards (If Applicable)*

- UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**TUV Product Services GmbH (TUVPSG)**

[Docket No. NRTL-1-2001]

*Withdrawn or Replaced Standards*

- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*

*Comparable Replacement Standards (If Applicable)*

- UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements

**TUV Rheinland of North America, Inc. (TUV)**

[Docket No. NRTL3-92]

*Withdrawn or Replaced Standards*

- UL 136 Pressure Cookers (not appropriate—no NRTL approval requirement)
- UL 1409 Low-Voltage Video Products Without Cathode-Ray-Tube Displays
- UL 1570 Fluorescent Lighting Fixtures
- UL 1571 Incandescent Lighting Fixtures
- UL 1572 High Intensity Discharge Lighting Fixtures
- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*
- UL 3121-1 Process Control Equipment
- UL 8730-1 Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 8730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 8730-2-4 Automatic Electrical Controls for Household and Similar

Use; Part 2: Particular Requirements for Thermal Motor Protectors for Motor Compressors or Hermetic and Semi-Hermetic Type\*\*

- UL 8730-2-8 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Water Valves

*Comparable Replacement Standards (If Applicable)*

- UL 1262 Laboratory Equipment (requested by NRTL)
- UL 1598 Luminaries (already included in NRTL's scope)
- UL 60730-1A Automatic Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 60730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls
- UL 61010A-1 Electrical Equipment For Laboratory Use; Part 1: General Requirements
- UL 61010C-1 Process Control Equipment

**Underwriters Laboratories Inc. (UL)**

[Docket No. NRTL4-93]

*Withdrawn or Replaced Standards*

- ANSI C37.71 Three Phase, Manually Operated Subsurface Load Interrupting Switches for Alternating-Current Systems
- ANSI C57.12.13 Liquid-Filled Transformers Used in Unit Installations including Unit Substations-Conformance Requirements
- ANSI C57.12.21 Pad Mounted Compartmental-Type Self-Cooled Single-Phase Distribution Transformers with High Voltage Bushings; 167 kVA and Smaller
- ANSI C57.12.27 Liquid Filled Distribution Transformers Used in Pad Mounted Installations, Including Unit Substations-Conformance Requirements
- ANSI C62.1 Gapped Silicon-Carbide Surge Arresters for AC Power Circuits
- ANSI Z21.10.2 Water Heaters-Sidearm Type Water Heaters
- ANSI Z21.11.1 Gas-Fired Room Heaters-Vented Room Heaters
- ANSI Z21.14 Approval Requirements for Industrial Gas Boilers
- ANSI Z21.16 Gas Unit Heaters
- ANSI Z21.29 Listing Requirements for Furnace Temperature Limit Controls and Fan Controls

- ANSI Z21.37 Approval Requirements for Dual Oven Type Combination Gas Ranges
- ANSI Z21.48 Gas-Fired Gravity and Fan Type Floor Furnaces
- ANSI Z21.49 Gas-Type Gravity and Fan Type Vented Wall Furnaces
- ANSI Z21.53 Gas-Fired Heavy Duty Forced Air Heaters
- ANSI Z21.55 Gas-Fired Sauna Heaters
- ANSI Z21.70 Earthquake Actuated Automatic Gas Shutoff Systems
- ANSI Z83.3 Gas Utilization Equipment in Large Boilers
- ANSI Z83.10 Separated Combustion System Central Furnaces
- ANSI Z83.17 Direct Gas Fired Door Heaters
- UL 198B Class H Fuses
- UL 198C High-Interrupting-Capacity Fuses, Current Limiting Type
- UL 198D Class K Fuses
- UL 198E Class R Fuses
- UL 198F Plug Fuses
- UL 198G Fuse for Supplementary Overcurrent Protection
- UL 198H Class T Fuses
- UL 198L D-C Fuses for Industrial Use
- ANSI/NEMA 250 Enclosures for Electrical Equipment
- UL 900 Air-Filter Units (not appropriate—no NRTL approval requirement)
- UL 910 Test for Flame-Propagation and Smoke-Density Values for Electrical and Optical-Fiber Cables Used in Spaces Transporting Environmental Air
- UL 1087 Molded-Case Switches
- UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*
- UL 1270 Radio Receivers, Audio Systems, and Accessories
- UL 1409 Low-Voltage Video Products Without Cathode-Ray-Tube Displays
- UL 1410 Television Receivers and High-Voltage Video Products
- UL 1570 Fluorescent Lighting Fixtures
- UL 1571 Incandescent Lighting Fixtures
- UL 1572 High Intensity Discharge Lighting Fixtures
- UL 3101-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3101-2-20 Electrical Equipment for Laboratory Use; Part 2: Laboratory Centrifuges Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*\*
- UL 3121-1 Process Control Equipment
- UL 8730-1 Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 8730-2-3 Automatic Electrical Controls for Household and Similar

- Use; Part 2: Particular Requirements for Thermal Motor Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 8730-2-4 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Motor Protectors for Motor Compressors or Hermetic and Semi-Hermetic Type\*\*
- UL 8730-2-6 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Pressure Sensing Controls Including Mechanical Requirements
- UL 8730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches
- UL 8730-2-8 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Water Valves
- UL 8730-2-9 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Temperature Sensing Controls\*\*
- UL 8730-2-14 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electric Actuators
- UL 60730-2-10 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically-Operated Motor Starting Relays
- UL 60730-2-11 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Energy Regulators
- UL 60730-2-12 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically-Operated Doors
- UL 60730-2-13 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Humidity Sensing Controls
- UL 60730-2-16 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level-Operating Controls of the Float Type for Household and Similar Applications
- Comparable Replacement Standards (If Applicable)*
- UL 1598 Luminaries (already included in NRTL's scope)
- UL 60730-1A Automatic Electrical Controls for Household and Similar Use; Part 1: General Requirements
- UL 60730-2-3 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Thermal Protectors for Ballasts for Tubular Fluorescent Lamps
- UL 60730-2-6 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Pressure Sensing Controls Including Mechanical Requirements
- UL 60730-2-7 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Timers and Time Switches
- UL 60730-2-10A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Motor Starting Relays
- UL 60730-2-11A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Energy Regulators
- UL 60730-2-12A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electrically Operated Door Locks
- UL 60730-2-13A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Humidity Sensing Controls
- UL 60730-2-14 Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Electric Actuators
- UL 60730-2-16A Automatic Electrical Controls for Household and Similar Use; Part 2: Particular Requirements for Automatic Electrical Water Level Controls
- UL 61010A-1 Electrical Equipment for Laboratory Use; Part 1: General Requirements
- UL 61010A-2-020 Electrical Equipment for Laboratory Use; Part 2: Particular Requirements for Laboratory Centrifuges
- UL 61010C-1 Process Control Equipment
- Wyle Laboratories, Inc. (WL)**  
[Docket No. NRTL1-93]
- Withdrawn or Replaced Standards*
- UL 198B Class H Fuses
- UL 198C High-Interrupting-Capacity Fuses, Current Limiting Type
- UL 198D Class K Fuses
- UL 198E Class R Fuses
- UL 198F Plug Fuses
- UL 198G Fuse for Supplementary Overcurrent Protection
- UL 198H Class T Fuses
- UL 198L D-C Fuses for Industrial Use
- UL 1087 Molded-Case Switches
- UL 1244 Electrical and Electronic Measuring and Testing Equipment\*\*
- UL 1570 Fluorescent Lighting Fixtures
- UL 1571 Incandescent Lighting Fixtures
- Comparable Replacement Standards (If Applicable)*
- UL 1598 Luminaries (already included in NRTL's scope)

UL 3111-1 Electrical Measuring and Test Equipment; Part 1: General Requirements\*

**\*Note on addition of UL 3111-1 to Wyle's scope:** This test standard is comparable to UL 1244, which is currently included in Wyle's scope of recognition. Wyle requested the addition of UL 3111-1 to its scope due to the pending withdrawal of UL 1244. However, as noted in the note below, UL 3111-1 is also due to be replaced. Therefore, the following note also applies to the addition of this standard.

**\*\*Note on test standard to be withdrawn:** This standard is due to be withdrawn by the standards development organization. However, as of the date of this notice, the OSHA NRTL Program staff has not determined whether or not there will be a replacement test standard. As a result, we will not remove this standard from the scope of the NRTL upon publication of this notice. Once we make the above determination, OSHA will replace or remove this standard, as appropriate. This note applies to the following test standards: UL 1244, UL 3111-1, UL 8730-2-4, and UL 8730-2-9.

In accordance with OSHA policy pertaining to recognition of replacement standards, the Agency only publishes one **Federal Register** notices to note the changes to the NRTL's scope of recognition. Changes to each NRTL's recognition are limited to those described in this notice. All other terms and conditions of each NRTL's recognition remain the same.

Signed at Washington, DC this 17th day of December, 2002.

**John L. Henshaw,**

*Assistant Secretary.*

[FR Doc. 03-123 Filed 1-3-03; 8:45 am]

**BILLING CODE 4510-26-P**

## DEPARTMENT OF LABOR

### Occupational Safety and Health Administration

[Docket No. NRTL-1-89]

#### Intertek Testing Services, NA, Inc., Application for Expansion of Recognition

**AGENCY:** Occupational Safety and Health Administration (OSHA), Labor.

**ACTION:** Notice.

**SUMMARY:** This notice announces the application of Intertek Testing Services, NA, Inc., for expansion of its recognition as a Nationally Recognized Testing Laboratory under 29 CFR 1910.7, and presents the Agency's preliminary finding. This preliminary finding does not constitute an interim or temporary approval of this application.