

including Appendices 1 and 2, dated September 26, 2013.

(l) Exception for Previously Replaced Fasteners

Replacement of all fractured and incorrectly oriented forward and aft fasteners, as specified in paragraph (i) or (k) of AD 2014–03–17, Amendment 39–17754 (79 FR 9389, February 19, 2014), if done before the effective date of this AD, is considered acceptable for compliance with the requirements of this AD.

(m) Exception to the Service Information

Where Bombardier Alert Service Bulletin A600–0763, Revision 02, dated December 9, 2014, including Appendices 1 and 2, dated September 26, 2013; and Bombardier Alert Service Bulletin A601–0627, Revision 02, dated December 9, 2014, including Appendices 1 and 2, dated September 26, 2013; specify to contact Bombardier for repair instructions, before further flight, repair using a method approved by the Manager, New York Aircraft Certification Office (ACO), FAA; or Transport Canada Civil Aviation (TCCA); or Bombardier's TCCA Design Approval Organization (DAO).

(n) Credit for Previous Actions

This paragraph provides credit for actions required by paragraphs (g), (h), and (i) of this AD, if those actions were performed before the effective date of this AD using the applicable service information identified in paragraphs (n)(1) through (n)(4) of this AD.

(1) Bombardier Alert Service Bulletin A600–0763, including Appendices 1 and 2, dated September 26, 2013, which was previously incorporated by reference on March 6, 2014 (79 FR 9389, February 19, 2014).

(2) Bombardier Alert Service Bulletin A600–0763, Revision 01, dated February 26, 2014, including Appendices 1 and 2, dated September 26, 2013, which is not incorporated by reference in this AD.

(3) Bombardier Alert Service Bulletin A601–0627, including Appendices 1 and 2, dated September 26, 2013, which was previously incorporated by reference on March 6, 2014 (79 FR 9389, February 19, 2014).

(4) Bombardier Alert Service Bulletin A601–0627, Revision 01, dated February 26, 2014, including Appendices 1 and 2, dated September 26, 2013, which is not incorporated by reference in this AD.

(o) Other FAA AD Provisions

The following provisions also apply to this AD:

(1) *Alternative Methods of Compliance (AMOCs)*: The Manager, New York ACO, ANE–170, FAA, has the authority to approve AMOCs for this AD, if requested using the procedures found in 14 CFR 39.19. In accordance with 14 CFR 39.19, send your request to your principal inspector or local Flight Standards District Office, as appropriate. If sending information directly to the ACO, send it to ATTN: Program Manager, Continuing Operational Safety, FAA, New York ACO, 1600 Stewart Avenue, Suite 410, Westbury, NY 11590; telephone 516–228–7300; fax 516–794–5531. Before

using any approved AMOC, notify your appropriate principal inspector, or lacking a principal inspector, the manager of the local flight standards district office/certificate holding district office. The AMOC approval letter must specifically reference this AD.

(2) *Contacting the Manufacturer*: For any requirement in this AD to obtain corrective actions from a manufacturer, the action must be accomplished using a method approved by the Manager, New York ACO, ANE–170, Engine and Propeller Directorate, FAA; or TCCA; or Bombardier, Inc.'s TCCA DAO. If approved by the DAO, the approval must include the DAO-authorized signature.

(p) Special Flight Permits

Special flight permits to operate the airplane to a location where the airplane can be repaired in accordance with sections 21.197 and 31.199 of the Federal Aviation Regulations (14 CFR 21.197 and 21.199) are not allowed.

(q) Related Information

(1) Refer to Mandatory Continuing Airworthiness Information (MCAI) Canadian Emergency Airworthiness Directive CF–2013–39R2, dated December 12, 2014, for related information. You may examine the MCAI on the Internet at <http://www.regulations.gov> by searching for and locating Docket No. FAA–2015–0082.

(2) Service information identified in this AD that is not incorporated by reference is available at the addresses specified in paragraphs (r)(3) and (r)(4) of this AD.

(r) Material Incorporated by Reference

(1) The Director of the Federal Register approved the incorporation by reference (IBR) of the service information listed in this paragraph under 5 U.S.C. 552(a) and 1 CFR part 51.

(2) You must use this service information as applicable to do the actions required by this AD, unless this AD specifies otherwise.

(i) Bombardier Alert Service Bulletin A600–0763, Revision 02, dated December 9, 2014, including Appendices 1 and 2, dated September 26, 2013.

(ii) Bombardier Alert Service Bulletin A601–0627, Revision 02, dated December 9, 2014, including Appendices 1 and 2, dated September 26, 2013.

(3) For service information identified in this AD, contact Bombardier, Inc., 400 Côte-Vertu Road West, Dorval, Québec H4S 1Y9, Canada; telephone 514–855–5000; fax 514–855–7401; email thd.crj@aero.bombardier.com; Internet <http://www.bombardier.com>.

(4) You may view this service information at the FAA, Transport Airplane Directorate, 1601 Lind Avenue SW., Renton, WA. For information on the availability of this material at the FAA, call 425–227–1221.

(5) You may view this service information that is incorporated by reference at the National Archives and Records Administration (NARA). For information on the availability of this material at NARA, call 202–741–6030, or go to: <http://www.archives.gov/federal-register/cfr/ibr-locations.html>.

Issued in Renton, Washington, on January 20, 2015.

Jeffrey E. Duven,

Manager, Transport Airplane Directorate, Aircraft Certification Service.

[FR Doc. 2015–01661 Filed 2–2–15; 8:45 am]

BILLING CODE 4910–13–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 870

[Docket No. FDA–2013–N–0234]

Effective Date of Requirement for Premarket Approval for Automated External Defibrillator Systems; Republication

AGENCY: Food and Drug Administration, HHS.

ACTION: Final order; republication.

SUMMARY: The Food and Drug Administration (FDA or the Agency) is republishing in its entirety a final order entitled “Effective Date of Requirement for Premarket Approval for Automated External Defibrillator” that published in the **Federal Register** on January 29, 2015 (80 FR 4783). FDA is republishing to correct an inadvertent omission of a comment regarding adverse tissue reaction as a risk to health and the Agency's response to that comment. The final order requires the filing of premarket approval applications (PMA) for automated external defibrillator (AED) systems, which consist of an AED and those AED accessories necessary for the AED to detect and interpret an electrocardiogram and deliver an electrical shock (e.g., pad electrodes, batteries, adapters, and hardware keys for pediatric use).

DATES: This order is effective on February 3, 2015.

FOR FURTHER INFORMATION CONTACT: Linda Ricci, Center for Devices and Radiological Health, 10903 New Hampshire Ave., Bldg. 66, rm. 1314, Silver Spring, MD 20993, 301–796–6325, linda.ricci@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background—Regulatory Authorities

The Federal Food, Drug, and Cosmetic Act (the FD&C Act), as amended by the Medical Device Amendments of 1976 (the 1976 amendments) (Pub. L. 94–295), the Safe Medical Devices Act of 1990 (Pub. L. 101–629), the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105–115), the Medical Device User Fee and Modernization Act

of 2002 (Pub. L. 107–250), the Medical Devices Technical Corrections Act (Pub. L. 108–214), the Food and Drug Administration Amendments Act of 2007 (Pub. L. 110–85), and the Food and Drug Administration Safety and Innovation Act (FDASIA) (Pub. L. 112–144), among other amendments, established a comprehensive system for the regulation of medical devices intended for human use. Section 513 of the FD&C Act (21 U.S.C. 360c) established three categories (classes) of devices, reflecting the regulatory controls needed to provide reasonable assurance of their safety and effectiveness. The three categories of devices are class I (general controls), class II (special controls), and class III (premarket approval).

Under section 513(d) of the FD&C Act, devices that were in commercial distribution before the enactment of the 1976 amendments, May 28, 1976 (generally referred to as “preamendments devices”), are classified after FDA has: (1) Received a recommendation from a device classification panel (an FDA advisory committee); (2) published the panel’s recommendation for comment, along with a proposed regulation classifying the device; and (3) published a final regulation classifying the device. FDA has classified most preamendments devices under these procedures.

Devices that were not in commercial distribution prior to May 28, 1976 (generally referred to as “postamendments devices”) are automatically classified by section 513(f) of the FD&C Act into class III without any FDA rulemaking process. Those devices remain in class III and require premarket approval unless, and until, the device is reclassified into class I or II or FDA issues an order finding the device to be substantially equivalent, in accordance with section 513(i) of the FD&C Act, to a predicate device that does not require premarket approval. The Agency determines whether new devices are substantially equivalent to predicate devices by means of premarket notification procedures in section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

A preamendments device that has been classified into class III and devices found substantially equivalent by means of premarket notification (510(k)) procedures to such a preamendments device or to a device within that type (both the preamendments and substantially equivalent devices are referred to as preamendments class III devices) may be marketed without submission of a premarket approval

application (PMA) until FDA issues a final order under section 515(b) of the FD&C Act (21 U.S.C. 360e(b)) requiring premarket approval or until the device is subsequently reclassified into class I or class II. Section 515(b)(1) of the FD&C Act directs FDA to issue an order requiring premarket approval for a preamendments class III device.

Although, under the FD&C Act, the manufacturer of a class III preamendments device may respond to the call for PMAs by filing a PMA or a notice of completion of a product development protocol (PDP), in practice, the option of filing a notice of completion of a PDP has not been used. For simplicity, although corresponding requirements for PDPs remain available to manufacturers in response to a final order under section 515(b) of the FD&C Act, this document will refer only to the requirement for the filing and receiving approval of a PMA.

On July 9, 2012, FDASIA was enacted. Section 608(a) of FDASIA (126 Stat. 1056) amended section 513(e) of the FD&C Act, changing the mechanism for reclassifying a device from rulemaking to an administrative order. Section 608(b) of FDASIA amended section 515(b) of the FD&C Act changing the mechanism for requiring premarket approval for a preamendments class III device from rulemaking to an administrative order.

Section 515(b)(1) of the FD&C Act sets forth the process for issuing a final order. Specifically, prior to the issuance of a final order requiring premarket approval for a preamendments class III device, the following must occur: (1) Publication of a proposed order in the **Federal Register**; (2) a meeting of a device classification panel described in section 513(b) of the FD&C Act; and (3) consideration of comments from all affected stakeholders, including patients, payers, and providers.

Section 515(b)(3) of the FD&C Act provides that FDA shall, after the close of the comment period on the proposed order, consideration of any comments received, and a meeting of a device classification panel described in section 513(b) of the FD&C Act, issue a final order to require premarket approval or publish a document terminating the proceeding together with the reasons for such termination.

A preamendments class III device may be commercially distributed without a PMA until 90 days after FDA issues a final order (a final rule issued under section 515(b) of the FD&C Act prior to the enactment of FDASIA is considered to be a final order for purposes of section 501(f) of the FD&C Act (21 U.S.C. 351(f))) requiring

premarket approval for the device, or 30 months after final classification of the device under section 513 of the FD&C Act, whichever is later. For AED systems, the later of these two time periods is the 90-day period. Therefore, section 501(f)(2)(B) of the FD&C Act requires that a PMA for such devices be filed within 90 days of the effective date of a final order. However, for the reasons discussed below, FDA does not intend to enforce compliance with the 90-day deadline for PMA submissions for currently marketed AEDs and those AED accessories identified in § 870.5310(a) (21 CFR 870.5310(a)) (see further discussion in section V, “Implementation Strategy”).

Also, a preamendments device subject to the order process under section 515(b) of the FD&C Act is not required to have an approved investigational device exemption (IDE) (see part 812 (21 CFR part 812)) contemporaneous with its interstate distribution until the date identified by FDA in the final order requiring the filing of a PMA for the device. At that time, an IDE is required only if a PMA has not been filed. If the manufacturer, importer, or other sponsor of the device submits an IDE application and FDA approves it, the device may be distributed for investigational use. If a PMA is not filed by the later of the two dates, and the device is not distributed for investigational use under an IDE, the device is deemed to be adulterated within the meaning of section 501(f)(1)(A) of the FD&C Act, and subject to seizure and condemnation under section 304 of the FD&C Act (21 U.S.C. 334) if its distribution continues. Other enforcement actions include, but are not limited to, the following: Shipment of devices in interstate commerce may be subject to injunction under section 302 of the FD&C Act (21 U.S.C. 332), and the individuals responsible for such shipment may be subject to prosecution under section 303 of the FD&C Act (21 U.S.C. 333). FDA requests that manufacturers take action to prevent the further use of devices for which no PMA has been filed.

II. Regulatory History of This Device

On January 25, 2011, the Circulatory System Devices Panel (“Panel”) recommended that AED systems be classified as class III devices and subject to premarket approval to provide reasonable assurance of the safety and effectiveness of the device (Ref. 1). The Panel recommended that AED systems be regulated as class III devices because, among other things, they are lifesaving devices. Furthermore, the problems identified in adverse events in the

medical device reporting systems and recalls related to AED systems indicated these devices require more regulatory oversight.

FDA published a proposed order to require PMAs for AED systems in the **Federal Register** of March 25, 2013 (78 FR 17890). FDA is now requiring PMAs for AED systems, which include AED accessories necessary for the functionality of the AED (e.g., pad electrodes, batteries, adapters, and hardware keys for pediatric use) (“necessary AED accessories”) (see section IV, “The Final Order”).

FDA received and has considered comments on the AED systems proposed order as discussed in section III of this document.

III. Public Comments in Response to the Proposed Order

In response to the March 25, 2013 (78 FR 17890) proposed order to maintain the class III classification and require premarket approval for AED systems, FDA received 66 comments and one petition for reclassification (see FDA–2013–N–0234–0002) (Ref. 2). The comments and the FDA’s responses to the comments are summarized below. Certain comments are grouped together under a single number because the subject matter of the comments is similar. The number assigned to each comment is purely for organizational purposes and does not signify the comment’s value or importance or the order in which it was submitted.

(*Comment 1*) Many comments indicated that AED systems have already been demonstrated to be safe and effective, and referenced literature and studies supporting the reliability of these devices and the value of AED systems in treating sudden cardiac arrest (SCA). The comments stated that PMAs and associated increased regulatory cost and review time is not warranted and would hinder innovation, increase device cost to consumers, and reduce availability of AED systems. The comments further stated that it is widely recognized that improvement in the survival rate from SCA is due in large part to widespread distribution of AED systems and expressed concern that requiring PMAs would limit availability of the devices.

(*Response 1*) FDA agrees that many currently marketed AEDs have been demonstrated to be effective in clinical use and, when designed and manufactured appropriately, AEDs can be safe and effective. However, FDA believes that there is insufficient information to determine that general and special controls would provide a reasonable assurance of the safety and

effectiveness of these devices, which are for a use in supporting or sustaining human life (see section 513(a)(1)(C) of the FD&C Act. Specifically, the postmarket information on AEDs supports increased regulatory review to ensure that device design and manufacturing practices provide a reasonable assurance of safety and effectiveness. FDA acknowledges that the PMA process may result in increased regulatory cost to manufacturers; however, FDA believes that device quality will improve, which will reduce costs associated with postmarket actions including recalls.

FDA also agrees that continued efforts to make safe and effective AED systems available is in the interest of public health, but disagrees that this call for PMAs will limit device availability. FDA believes that many manufacturers of currently marketed AEDs already have, or can reasonably obtain, the necessary data to support a PMA, and hence expects AED distribution to continue to meet demand. Also, for the reasons discussed below, FDA does not intend to enforce compliance with the 90-day deadline for submission of PMAs for currently marketed AEDs and necessary AED accessories (for further discussion see section V, “Implementation Strategy”).

At the January 2011 Panel meeting, the Panel discussed the impact of FDA regulatory scrutiny on innovation. Various Panel members agreed that the appropriate focus should be on assuring reliability of AEDs and that there was no evidence presented to indicate that a call for PMAs would unduly hinder device innovation (Ref. 1). FDA notes that previous significant innovations for AED systems (e.g., new defibrillation waveforms) have been supported by clinical evidence in the 510(k) process and that under the PMA process this clinical evidence is not expected to significantly change. As was mentioned in the proposed order, FDA anticipates that many AED manufacturers already have sufficient clinical evidence to support a PMA.

(*Comment 2*) Several comments noted that AED system failures are often the result of use error or improper maintenance (e.g., expired batteries/pads, periodic checks not performed, etc.) and not of system failure or malfunction. The comments stated that efforts should be devoted to ensuring appropriate public awareness, training (particularly for lay users), and maintenance to address these issues as opposed to increasing premarket regulatory review. One comment stated that the proposed order should not be finalized until all stakeholders, not only

device manufacturers, are engaged in an integrated approach to increase the likelihood that AED systems will be available and functional when needed.

(*Response 2*) FDA agrees that AED system training and maintenance are important to help ensure AED system availability and proper use and also believes manufacturers and users are in the best position to develop and implement training and maintenance materials. FDA supports ongoing discussions and efforts to improve training and maintenance, but disagrees that these activities should delay finalizing the requirement for PMAs for these devices. Although we recognize that there have been some medical device reports (MDRs) associated with use errors, the focus of FDA’s review of MDRs and recalls of AED systems has been related to problems with the quality of these devices as related to device design and manufacture and FDA continues to believe that requiring PMAs is appropriate.

(*Comment 3*) Several comments stated that special controls, including performance testing to industry standards, device labeling, guidance documents, human factors analysis and design, summary of field actions and mitigations to address Quality System (QS) concerns, risk management, and post-market surveillance were sufficient to regulate AED systems as class II devices under the existing 510(k) regulatory regime. One comment indicated that several of the regulatory controls identified by FDA as consistent with PMA requirements—such as premarket inspections, review of changes that could significantly affect the safety or effectiveness of the device, and postmarket surveillance—could also be conducted under the 510(k) regime. Other comments supported FDA’s proposal to maintain the devices in class III and agreed that the manufacturing controls, premarket review requirements, and assessment of lay use are best managed under the PMA process.

(*Response 3*) FDA disagrees that there is sufficient information to determine that general and special controls would provide a reasonable assurance of safety and effectiveness of these devices given safety concerns related to the manufacturing processes and design changes, problems which FDA considered in determining that PMAs are warranted (see section 513(a)(1)(C) of the FD&C Act. FDA does not generally conduct preclearance inspections under the 510(k) process because such information is not required in a 510(k) submission under the FD&C Act or FDA regulations.

Further, under section 513(f)(5) of the FD&C Act, FDA may not withhold a 510(k) because of a failure to comply with any provision of this Act unrelated to a substantial equivalence decision, including a finding that the facility in which the device is manufactured is not in compliance with good manufacturing requirements as set forth in regulations of the Secretary under section 520(f) (other than a finding that there is a substantial likelihood that the failure to comply with such regulations will potentially present a serious risk to human health). In contrast, under section 515(c)(1)(C) of the FD&C Act, a PMA must include a full description of the methods used in, and the facilities and controls used for, the manufacturing, processing, and when relevant, packing and installation of, such device. Moreover, many of the design and manufacturing changes that have led to AED system recalls were not required to be reported to FDA under the 510(k) process. If these changes had been reported prior to implementation, as would be required in the PMA regime, these recalls may have been avoided. FDA continues to believe that the necessary regulatory controls for AED systems are consistent with the PMA review process, and that the 510(k) process does not provide sufficient regulatory oversight for these devices.

Similarly, FDA's oversight of postmarket changes to devices is very different in the 510(k) context as compared to the PMA context. Under § 807.81, FDA requires 510(k)s for a change to a device only when the change could significantly affect the safety or effectiveness of the device, *e.g.*, a significant change or modification in design, material, chemical composition, energy source, or manufacturing process. In contrast, under 21 CFR 814.39, FDA requires PMA supplements (including 30-day notices) for any change to a PMA-approved device that affects safety or effectiveness. These differences in authorities, among the other reasons discussed previously, warrant regulation of AEDs in class III.

(Comment 4) A few comments indicated that existing AED and AED accessory manufacturers are already subject to the QS regulation (21 CFR part 820) and manufacturing quality would not be measurably improved as a result of requiring PMAs. One comment noted that specific expectations under the QS regulation for design controls, purchasing controls, and other issues identified by FDA as problematic for AEDs could be addressed by special controls and other regulations, and AEDs could remain in class II. One comment further stated that such

concerns could be managed via postmarket controls, which are available under the 510(k) regime, such as submission of a summary of recent field actions and related design mitigations.

(Response 4) FDA disagrees with the comments. FDA acknowledges that AED and AED accessory manufacturers are already subject to the QS regulation and that QS requirements result in rigorously designed and manufactured devices and resultant quality improvements. By requiring premarket review of QS processes as well as device changes for AEDs, FDA believes the PMA process will provide a reasonable assurance of safety and effectiveness (see Response 3).

(Comment 5) One comment stated that certain AED accessories, specifically electrodes, cables, and adapters, are well-understood devices and that their classification into class III is not warranted. The comment stated that these accessories could be adequately regulated in class II with special controls, as is already the case when these accessories are used with manual defibrillators. The comment recommended special controls, including the following: performance testing, usability evaluation, labeling, biocompatibility, and readiness for use. Two comments stated that because AED accessories often have identical designs and the same intended use as accessories used with class II manual defibrillators, FDA should not perform duplicative reviews under both the 510(k) and PMA regimes and that PMA review should be required only when use of the accessory with an AED results in a change in intended use or design.

(Response 5) Accessories necessary for an AED to detect and interpret an electrocardiogram and deliver an electrical shock (*e.g.*, battery, pad electrode, adapter, and hardware keys for pediatric use) are necessary for AED system functionality. Failure of these necessary accessories leads to the same negative outcomes as a failure of the AED itself; *e.g.*, an AED not ready for use because of a faulty battery is unable to detect heart rhythm abnormalities and/or deliver a defibrillation shock to a victim of SCA. FDA's review of adverse events and recalls has shown that problems with AED accessories have occurred during clinical use. As such, FDA continues to believe that the same regulatory oversight is warranted for certain critical accessories (*i.e.*, batteries, pad electrodes, adapters, and hardware keys for pediatric use) as for the AEDs with which they are used. As discussed in the response to Comment 3, FDA does not believe that adequate regulatory controls are available under

the 510(k) process, and hence PMAs are necessary to provide a reasonable assurance of safety and effectiveness.

(Comment 6) Several comments questioned the validity of FDA's data regarding adverse events associated with AED failures. One comment noted that FDA provided no data in the proposed order on the frequency of adverse events or relationship of number of events to total distribution and use of AEDs. The comment requested additional information from FDA to support the validity of the MDR analysis presented at the 2011 Panel and relied upon to support the proposed order. A few comments presented alternate analyses of MDR data that suggested that MDRs for AEDs are not increasing. One comment presented an analysis that showed no statistically significant increase in the rate of adverse event reports over the time period of 2007 to 2011. Two comments stated that a majority of AED MDRs reported to FDA resulted from self-test errors—which are reported as malfunction MDRs because they could cause or contribute to a death or serious injury but do not represent device failures in clinical use. The comments contended that any analysis of MDRs should focus instead on actual use adverse events, which would represent a small subset of the overall MDRs. One comment stated that self-test related events are representative of an effective design risk mitigation strategy being employed for AEDs and that because AEDs are often in standby for a large percentage of time, self-test detection of problems before use should not be included in the overall assessment of the benefit-risk profile for AEDs. Two comments requested further guidance from FDA on MDR reporting expectations for AEDs.

(Response 6) Although FDA requires manufacturers to submit an MDR when their device may have caused or contributed to a death, serious injury, or in certain situations when their device has malfunctioned, FDA acknowledges that there are limitations on the review of MDR data, including the fact that FDA typically does not have complete information on the number of devices in distribution from which to calculate adverse events rates. These limitations were discussed at the 2011 Panel meeting. FDA has previously stated that fatality statistics and injury statistics from MDRs should be considered in light of underreporting (58 FR 61952 at 61972, November 23, 1993). In addition, FDA notes that the evaluation of MDR data for AEDs was focused on manufacturing and design concerns and was not aimed at developing specific

failure rates. Moreover, FDA believes that the large number of devices in distribution and the life-saving nature of these devices combined with the steady rate of MDRs support a call for PMAs to help ensure that these devices are adequately designed and manufactured so that they are available when needed.

FDA disagrees that evaluation of adverse events should focus only on those events that occur during clinical use. Although some distributed AEDs may seldom be used, this does not reduce the importance that they are safe and effective when needed. FDA acknowledges the importance of AED self-test features and recognizes that many self-test failures are not indicative of issues with overall device quality. FDA believes, however, that some self-test failures signal significant quality problems arising from device design or manufacturing issues and are appropriately considered as adverse events if recurrence of such failures could, for example, render the device unavailable for use when needed. FDA also recognizes that some MDRs may eventually be found to be the result of problems not associated with the device; however, this concern is applicable to all devices subject to adverse event reporting requirements and FDA does not believe such reports unduly influence overall reporting numbers.

FDA also notes that our review of available information, as presented at the January 2011 Panel meeting, included data on voluntary corrections and removals (*i.e.*, “recalls”) of AEDs pursuant to section 519(g) of the FD&C Act (21 U.S.C. 360i(g)). Recalls are conducted “(A) to reduce a risk to health posed by the device, or (B) to remedy a violation of this Act caused by the device which may present a risk to health,” and as such may reflect safety concerns for AEDs (section 519(g)(1) of FD&C Act (21 U.S.C. 360i(g)(1))). Since the January 2011 Panel meeting, over 40 additional class I or class II recalls have been conducted by AED manufacturers and have impacted over 2 million distributed AEDs (Ref. 3). The root cause of these recalls has been attributed to a variety of causes, with design controls, purchasing controls, and receiving acceptance activities being the most common. FDA continues to believe that the recall data reinforces the overall conclusion regarding the inadequacy of regulatory controls for AED systems under the 510(k) process.

Additional guidance on MDR requirements for AEDs is beyond the scope of this document; however, FDA intends to continue efforts to clarify medical device reporting expectations

and manufacturers who have questions regarding their reporting obligations should contact FDA.

(*Comment 7*) FDA received a recommendation regarding including adverse tissue reaction as a risk to health when using AEDs, and a recommendation to require biocompatibility testing as a special control to mitigate the risk, specifically by ensuring the biocompatibility of the patient-contacting materials. The patient-contacting materials of the device may produce local adverse tissue effects, such as skin rash or irritation. Device materials that are not biocompatible may either directly or through the release of their material constituents produce adverse local or systemic effects. Although medical devices may have myriad biocompatibility issues, the biocompatibility concerns from AEDs are likely limited to skin reactions from contact with the materials from which the pad electrode is made.

(*Response 7*) In the proposed order published in the **Federal Register** (78 FR 17890, March 25, 2013), FDA did not identify adverse tissue reaction as a risk associated with AEDs. However, FDA agrees that adverse tissue reaction is a risk to health for this device. For all of the reasons identified in the proposed order and this document, however, FDA believes that there is insufficient information to determine that general and special controls would provide a reasonable assurance of the safety and effectiveness of these devices. The risk of adverse tissue reactions, as well as the other risks posed by these devices, will be addressed during the premarket approval process of these devices.

(*Comment 8*) Several comments responded to FDA’s request for feedback regarding whether 15 months is sufficient to allow companies to collect information necessary to support submission of a PMA. Two comments stated that this issue was dependent on the data expected by FDA and that FDA should provide more guidance in this respect. One comment requested clarification on what clinical data is known to FDA that would support a PMA because it is critical that AED manufacturers understand the type and amount of data that will be required. One comment stated that it is unclear what FDA’s expectations would be for clinical trials of new AEDs or the need for clinical trials for AED accessories given available less burdensome methods for obtaining performance data on accessories. Another comment requested clarification on whether AED manufacturers would be expected to re-test and re-validate older AED models to

currently recognized standards. One comment requested clarification on when marketing materials for AEDs would need to comply with 21 CFR 801.109.

One comment suggested that the 15-month period should be extended to 30 months, which the commenter claimed would be consistent with section 501(f)(2)(B) of the FD&C Act. One commenter requested clarification regarding whether the 15 months started at the 90th day after a final order was issued and another comment indicated that 15 months should be sufficient, but that the 15 months should not include FDA’s 180-day PMA review time. One comment suggested that FDA require PMAs 90 days after the final order.

(*Response 8*) The data required to support premarket approval will vary by device and the specific data requirements. FDA is aware of clinical study information that can be leveraged for AEDs from both published studies and clinical data previously submitted to FDA under the 510(k) process, and, as was stated in the proposed order, FDA believes that many AED accessories “may need to submit non-clinical performance testing with confirmatory animal studies in order to support independent PMA approval” (78 FR 17890 at 17894, March 25, 2013). Performance testing of AEDs must be provided in a PMA to support a reasonable assurance of safety and effectiveness. Although retesting older AED models to currently recognized standards is one way to meet the performance testing requirements, compliance with such standards is voluntary and manufacturers may submit a justification for how other testing conducted on their devices provides equivalent assurances of safety and effectiveness. FDA encourages manufacturers to proactively engage FDA via the pre-submission process to discuss the specific data needed for their PMAs (Ref. 4). FDA notes that existing prescription AEDs are already subject to 21 CFR 801.109, and will remain so after this call for PMAs. FDA review of AED PMAs will include review of the associated AED labeling to ensure such device labeling complies with regulatory requirements.

FDA notes that the 30 months discussed in section 501(f)(2)(B) of the FD&C Act references the date from initial classification of a device into class III. AEDs have been classified as class III for more than 30 months, and hence this statutory provision has expired. FDA also acknowledges that it is in the interest of public health to ensure the availability of AEDs because they are life-saving devices and their

clinical use is well-established. After consideration of the comments, FDA continues to believe that the proposed 15 months for filing a PMA (Ref. 5) strikes an appropriate balance between the need to ensure continued availability of AEDs for the public health reasons stated previously and the implementation of PMA requirements to ensure the safety and effectiveness of AEDs.

For currently marketed AEDs, FDA does not intend to enforce compliance with the 90-day deadline by which PMAs must be submitted for 15 months after that deadline (*i.e.*, 18 months after the effective date of the final order), as long as a notice of intent to file a PMA is submitted within 90 days of the effective date of the final order (see section V, "Implementation Strategy"). Even if a notice of intent and PMA are submitted by these dates, manufacturers must cease distribution of devices upon receiving a not approvable or denial decision rendered on a PMA. To resume distribution, these manufacturers must receive PMA approval for their devices.

Moreover, for currently marketed necessary AED accessories, FDA does not intend to enforce compliance with the 90-day deadline by which PMAs must be submitted for 57 months after that deadline (*i.e.*, 5 years after the effective date of the final order) (see section V, "Implementation Strategy"). Continued availability of necessary AED accessories, including consumable accessory items (*e.g.*, pad electrodes) and accessories with limited useful life (*e.g.*, batteries), is critical to ensuring the safety and efficacy of currently marketed AEDs during the time while PMAs for those AEDs are being pursued. In addition, the continued availability of necessary accessories for "legacy devices"—individual AEDs that have been distributed and are currently in use (*e.g.*, in public facilities, etc.) and for which the manufacturer is not seeking PMA approval for that AED model—ensures the availability of functional legacy AEDs until they are replaced with PMA-approved AEDs.

(*Comment 9*) One commenter stated that FDA did not have a legal basis for continuing with finalization of a call for PMAs for AED systems because FDA failed to convene a panel as is required under FDASIA prior to issuing a final order. The commenter stated that FDA may not rely on the 2011 pre-FDASIA Panel because that Panel meeting was related to reclassifications under section 515(i) of the FD&C Act and not related to calls for PMAs under section 515(b). The commenter further contended that the 2011 Panel neither considered new information contained in a

reclassification petition submitted to FDA in 2009 nor adequately discussed the appropriateness of class II special controls.

(*Response 9*) FDA disagrees with the comment that FDA does not have a legal basis to finalize an order calling for PMAs for AED systems. Pursuant to FDASIA, the amendments to section 515(b) of the FD&C Act require, in relevant part, that issuance of an administrative order calling for PMAs for a preamendments device be preceded by a proposed order and a meeting of a classification panel. As amended, this section of the FD&C Act does not prescribe when these two events (the panel and proposed order) must occur in relation to each other. More importantly, FDA believes that the Panel's deliberations and recommendations remain relevant and fully satisfy the requirements in section 515(b) of FD&C Act.

FDA disagrees with the comment that the Panel did not consider new information contained in the 2009 reclassification petition. A representative from the petitioner was present at the meeting and provided comments on the reclassification petition during the Panel meeting (Ref. 1). In addition, the petitioner was given an opportunity to explain the petitioner's reasons for why AEDs should be class II devices, including a discussion of the special controls described in the reclassification petition. Therefore, the Panel heard the petitioner's arguments and these arguments were available for the Panel's consideration when it made its recommendation.

(*Comment 10*) One commenter objected to FDA's use of the term "diagnose" in the proposed order to describe the functionality of AEDs (78 FR 17890 at 17893, March 25, 2013), and stated that AEDs sense shockable rhythms and are not diagnostic devices.

(*Response 10*) FDA disagrees that these devices do not perform diagnostic functions. AEDs analyze and interpret ECG data to produce an assessment as to whether a shock should be delivered; while FDA does believe that AEDs have diagnostic functions, we note that the regulatory identification for the device in § 870.5310(a), as finalized in the order, does not use the term diagnose, and instead describes the function of the device as "analyzes" and "interprets."

(*Comment 11*) One commenter stated that FDA's proposal to allow manufacturers to "bundle" several AED models under a single PMA is inconsistent with the PMA regulatory paradigm, which relies on a device-by-device assessment. The comment points

to FDA's guidance on bundling, which states that "[g]enerally, [manufacturers] should not bundle differing generic device types in a single PMA submission because of the substantially different pre-clinical and clinical data needed to support each of the devices" (Ref. 6).

(*Response 11*) FDA disagrees with the comment. Different AED models can be included in one PMA if they are the same generic device type. Because shock advisory algorithms and defibrillation waveforms will likely be common across various models from a given manufacturer of devices, FDA expects the clinical data needed to support devices within an appropriately bundled AED PMA to be the same. However, because of the differences in device labeling and user requirements between professional and lay use devices, FDA continues to believe that separate PMAs should be submitted for a manufacturer's professional use versus lay use devices. FDA believes this approach is least burdensome to manufacturers and is consistent with the bundling guidance, which states that "[b]undling is appropriate for devices that present scientific and regulatory issues that can most efficiently be addressed during one review" (Ref. 6).

(*Comment 12*) One comment requested clarification on whether separate PMAs are required for AEDs and the associated AED accessories when a company manufactures both for use together. Two comments requested additional clarification on whether accessories not specified in the proposed order (such as electrocardiograph modules and electrodes, training pads/batteries, protective carrying cases, Bluetooth modules, hardware keys or specialized pads to reduce energy for pediatric use, self-testers, SpO₂/blood pressure monitoring devices, cardiopulmonary resuscitation (CPR) aids, medical device data systems (MDDS), etc.) would require PMAs. One comment suggested that AED accessories that are already 510(k) cleared should not be subject to premarket approval by virtue of being used with an AED.

(*Response 12*) In response to this comment, FDA has revised the identification language to clarify that AED accessories regulated under § 870.5310 are those accessories necessary for the AED to detect and interpret an electrocardiogram and deliver an electrical shock (*e.g.*, battery, pad electrode, adapter, and hardware keys for pediatric use). Manufacturers of accessory devices that are not addressed by the final order and are not already

the subject of an existing classification regulation should contact FDA.

Under the final order, manufacturers must submit PMAs for accessories that are necessary for operation of the AED system (e.g., accessories necessary to allow the AED to detect or interpret an electrocardiogram or deliver a defibrillation shock). These AED accessories include batteries, pad electrodes (including reduced energy pads for pediatric use), adapters, and hardware keys for pediatric use. In response to this comment, FDA has added "hardware keys for pediatric use" to the identification. Necessary AED accessories that are for use with a specific AED should be included in that PMA for the AED system as a whole. Alternatively, necessary AED accessories, including those manufactured by a third party, may be submitted in their own PMAs.

Accessories that are not necessary for the functionality of the AED are not addressed by the final order. Currently marketed AED accessories that are not addressed by the final order, such as SpO₂/blood pressure monitoring devices, ECG modules and testers, CPR aids, and MDDS, may be subject to other regulations and will continue to be subject to those existing regulations. Training accessories such as training pads and batteries for training-only AEDs are not currently subject to any additional regulations, and will not become so as a result of the final order.

(Comment 13) One comment requested clarification regarding AEDs being considered adulterated 90 days after the effective date of a final order in the absence of submission of a statement of intent to submit a PMA or the submission of a full PMA. The comment questioned whether devices legally distributed prior to the 90th day could remain in distribution. Another comment requested clarification on whether manufacturers could continue to provide consumable accessory items (such as batteries and pads) for previously distributed devices even if a PMA will not be submitted for that AED model. Two comments requested clarification on how and whether manufacturers would be allowed to distribute components required for field servicing of a device, including refurbished replacement devices, before PMAs are submitted for the devices.

(Response 13) Under the final order (see section IV, "the Final Order") and section 501(f)(2)(B) of the FD&C Act, PMAs must be submitted within 90 days after the effective date of the final order for currently marketed AED systems. If a PMA is not submitted, the devices are adulterated. However, for the reasons

discussed previously, for currently marketed AEDs, FDA does not intend to enforce compliance with the 90-day deadline by which PMAs must be submitted for 15 months after that deadline (i.e., 18 months after the effective date of the final order), as long as a notice of intent to file a PMA is submitted within 90 days of the effective date of the final order (see section V, "Implementation Strategy"). For currently marketed necessary AED accessories, FDA also does not intend to enforce compliance with the 90-day deadline by which PMAs must be submitted for 57 months after that deadline (i.e., 5 years after the effective date of the final order) (see section V, "Implementation Strategy"). This intention applies to necessary AED accessories regardless of whether a PMA is being or has been sought for the AED model.

(Comment 14) One comment indicated that premarket review of medical devices such as AEDs should include review of the software embedded into the devices, including review of software verification and validation documentation. The comment noted that such review should also occur for software modifications and software developed for maintenance of the devices, including self-test functions. The comment relayed the importance of having reviewers with adequate training, expertise, and experience.

(Response 14) FDA agrees with the comment. Review of AEDs under the 510(k) process has included a detailed review of software documentation supporting premarket submissions by appropriately trained and experienced FDA reviewers. The PMA review will also involve a review of software documentation and will be conducted by trained and experienced FDA reviewers.

(Comment 15) One comment suggested an alternative regulatory approach whereby AEDs for medical professional use be reclassified into class II and public access defibrillators used by laypersons remain in class III with PMAs required. The comment stated that professional use devices have advanced functionality and are operated by skilled and trained professionals, which lessens the chance of human factor errors and increases the likelihood that the user will be able to recognize and troubleshoot any malfunctions. The comment stated that such users can rely on past experience and other means of attempting to rescue a patient, whereas lay users are often fully reliant on the AED. Two comments also indicated that professional use

devices are typically manual defibrillators with additional functionality, including AED, and that the proposed order would create an inconsistent system whereby the same hardware if used only for manual defibrillation would be class II, but by virtue of configuring the device to include AED functionality would become a PMA class III product.

(Response 15) FDA disagrees with the comments and believes that the submission of PMAs is warranted regardless of the intended user of the device. FDA does not believe that there is sufficient information to establish special controls to provide a reasonable assurance of safety and effectiveness of the device regardless of the training of AED users.

FDA recognizes that some manual defibrillators and AEDs share common hardware and software platforms, and hence devices with similar or identical platforms may receive different regulatory review based on the configuration. For the reasons previously stated, however, FDA continues to believe AED systems should be class III devices. FDA also notes that the performance and other data needed to support safety and effectiveness for hardware and software platforms for both types of devices would be nearly identical; the difference would be related to the amount of information that must be submitted to FDA. For a PMA, more information on the design controls process is required to be submitted whereas for a 510(k) submission, some information may not need to be submitted and instead can reside within the company's overall quality system and associated design documentation. Such situations of different regulatory processes have occurred in other product areas including contact lenses (daily-wear are typically class II, whereas extended wear are class III) and ablation devices (general surgical use are class II, whereas use for treatment of atrial fibrillation is class III), and FDA does not believe this changes the overall rationale supporting the need for PMAs.

(Comment 16) Two comments noted that there are numerous companies that refurbish and/or resell AEDs. The comments requested that FDA include AED resellers and refurbishers in their consideration of regulatory strategy.

(Response 16) Regardless of the supplier, the introduction or delivery for introduction into interstate commerce of any device that is adulterated is a prohibited act under section 301 of the FD&C Act (21 U.S.C. 331) (see Comment 12). FDA encourages refurbishers and resellers who have

questions about the continued distribution of AEDs to contact FDA via the pre-submission process.

(Comment 17) One comment requested clarification of the process for modifications of currently marketed AEDs (and notifying FDA of such modifications) during the 90-day period after the final order is issued. One comment stated that given the nature of commercial, electrical and mechanical components used in AEDs, an extended transition period without the ability to implement changes would not be tenable and would result in unavailability of devices. One comment requested clarification on 510(k) submissions accepted for review, but for which no decision had been rendered, prior to the effective date of a final order calling for PMAs.

(Response 17) Under § 870.5310, as amended, all new AED and necessary AED accessories must have an approved PMA in effect before being placed in commercial distribution. We recommend that manufacturers of currently marketed AEDs contact FDA regarding implementation of any changes necessary for their AEDs in order to address safety concerns or to support ongoing distribution while PMA approval is being sought. FDA understands that issues may arise relating to part obsolescence or changes necessary to reduce a risk to health posed by a currently marketed AED that is not functioning properly.

All other changes need to be accounted for in a PMA. Moreover, all new AED and necessary AED accessories must have an approved PMA in effect before being placed in commercial distribution.

(Comment 18) One comment objected to the comparisons made by FDA at the 2011 Panel meeting between implanted cardioverter defibrillators (ICDs) (PMA devices) and AEDs. The comment noted the number of commercial components (e.g., components supplied to multiple industries for a variety of uses) in order to maintain affordable price-points for AEDs and the potentially burdensome PMA supplements that would be necessary to support incremental changes in manufacturing for AEDs. The comment further contended that purchased component-related recalls for AEDs have largely been a result of latent component failures and that FDA's examples at the 2011 Panel meeting of QS concerns related to changes to purchased components or device design would not have been averted by the stricter premarket regulatory oversight via PMA supplements.

(Response 18) FDA acknowledges that more stringent regulatory oversight via

the PMA process may not completely eliminate AED recalls. FDA also recognizes that AEDs typically contain commercial components and manufacturers will need to submit PMA supplements for changes to these components, as well as changes to suppliers and manufacturing processes. Use of commercial components in PMA devices is not uncommon and changes at the component level may have led to some AED recalls and adverse events, providing further support for increased regulatory review. FDA continues to believe that these failures and the need for careful consideration and adequate verification and validation of such changes support more rigorous review under the PMA process.

(Comment 19) One comment requested clarification on activities during the time after a notice of intent to file is submitted, including whether FDA will place additional postmarket approval requirements on previously distributed products as allowed under 21 CFR 814.82. The comment further asked whether IDEs would be required for design changes (e.g., would an IDE be required to conduct human factors/usability studies).

(Response 19) FDA will consider the need for postapproval requirements in the context of each manufacturer's PMA submission and the devices in distribution. FDA does not intend to exempt manufacturers from the IDE requirements and hence any study which meets the IDE requirements must be conducted in accordance with the requirements of 21 CFR parts 50 and 812. There will be no extended period for filing an IDE and studies may not be initiated without appropriate IDE approvals. Manufacturers who have questions regarding whether an IDE is needed for a particular AED study are encouraged to interact with FDA via the pre-submission process.

IV. The Final Order

FDA is adopting its findings as published in the preamble of the proposed order (78 FR 17890, March 25, 2013), with the addition of adverse tissue reaction as being identified as a risk to health in this final order, and is issuing this final order to require the filing of a PMA for AED systems under 515(b) of the FD&C Act. An AED system consists of an AED and those accessories necessary for the AED to detect and interpret an electrocardiogram and deliver an electrical shock (e.g., battery, pad electrode, and hardware keys for pediatric use). An AED system analyzes the patient's electrocardiogram, interprets the cardiac

rhythm, and automatically delivers an electrical shock (fully automated AED), or advises the user to deliver the shock (semi-automated or shock advisory AED) to treat ventricular fibrillation or pulseless ventricular tachycardia. Under section 515(b)(1)(A) of the FD&C Act, PMAs for AED systems are required to be filed on or before 90 days after the effective date of a final order. This final order will revise 21 CFR part 870.

V. Implementation Strategy

Based on comments on the proposed order, we are clarifying FDA's intentions regarding enforcing compliance with the final order (see section IV, "The Final Order") and section 501(f)(2)(B) of the FD&C Act.

A. Currently Marketed AEDs

Under the final order and section 501(f)(2)(B) of the FD&C Act, PMAs for currently marketed AEDs are required to be filed on or before 90 days after the effective date of a final order. However, for currently marketed AEDs, FDA does not intend to enforce compliance with this 90-day deadline for 15 months after that deadline (i.e., 18 months after the effective date of the final order), as long as notice of intent to file a PMA is submitted within 90 days of the effective date of the final order. The notification of the intent to file a PMA submission must include a list of all model numbers for which a manufacturer plans to seek marketing approval through a PMA.

In conducting any clinical studies, AEDs may be distributed for investigational use if the requirements of the IDE regulations (21 CFR part 812) are met. There will be no extended period for filing an IDE nor exemption from IDE requirements, and studies may not be initiated without appropriate IDE approvals, where necessary.

B. Currently Marketed Necessary AED Accessories

Under the final order and section 501(f)(2)(B), PMAs for currently marketed necessary AED accessories are required to be filed on or before 90 days after the effective date of this final order. However, for currently marketed necessary AED accessories, FDA does not intend to enforce compliance with this 90-day deadline for 57 months after the deadline (i.e., 5 years after the effective date of the final order). Currently marketed necessary AED accessory manufacturers are not required to file an intent-to-file by the 90-day deadline.

After the effective date of the final order, new AEDs and necessary AED accessories must have approved PMAs

to be legally marketed. The following tables show the regulatory timetable for currently marketed AEDs and necessary AED accessories.

TABLE 1—CURRENTLY MARKETED AEDS

	Timetable for which FDA does not intend to enforce compliance (time after effective date of order)	Distribution period (time after effective date of order)
Intent to File a PMA	90 days	<i>Devices included in an intent to file:</i> 18 months. <i>Devices not included in intent to file:</i> 90 days.
File a PMA	18 months	Until a not approvable decision or denial decision is issued; can continue distribution if an approval order is issued.

TABLE 2—CURRENTLY MARKETED NECESSARY AED ACCESSORIES

	Timetable for which FDA does not intend to enforce compliance (time after effective date of order)	Distribution period (time after effective date of order)
Intent to File a PMA	N/A	N/A.
File a PMA	60 months	Until a not approvable decision or denial decision is issued; can continue distribution if an approval order is issued.

VI. Environmental Impact

The Agency has determined under 21 CFR 25.30(h) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

VII. Paperwork Reduction Act of 1995

The final order refers to previously approved collections of information found in FDA regulations. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3520). The collections of information in 21 CFR part 812 have been approved under OMB control number 0910–0078; the collections of information in 21 CFR part 807, subpart E, have been approved under OMB control number 0910–0120; the collections of information in 21 CFR part 814, subpart B, have been approved under OMB control number 0910–0231; and the collections of information under 21 CFR part 801 have been approved under OMB control number 0910–0485.

VIII. Codification of Orders

Prior to the amendments by FDASIA, section 515(b) of the FD&C Act provided for FDA to issue regulations to require PMA approval for preamendments devices or devices found substantially equivalent to preamendments devices. Section 515(b) of the FD&C Act, as amended by FDASIA, provides for FDA to require PMA approval for such devices by issuing a final order, following the issuance of a proposed order in the **Federal Register**. FDA will continue to codify the requirement for a

PMA approval in the Code of Federal Regulations. Therefore, under section 515(b)(1)(A) of the FD&C Act, as amended by FDASIA, in this final order, we are requiring PMA approval for AED systems and we are making the language in § 870.5310 consistent with the final version of this order.

IX. References

The following references have been placed on display in the Division of Dockets Management (see **ADDRESSES**) and may be seen by interested persons between 9 a.m. and 4 p.m., Monday through Friday. (FDA has verified the Web site addresses, but FDA is not responsible for any subsequent changes to the Web sites after this document publishes in the **Federal Register**.)

1. Meeting Materials for 515(i) Regulatory Classification of Automated External Defibrillator Systems, January 25, 2011, available at <http://www.fda.gov/AdvisoryCommittees/CommitteesMeetingMaterials/MedicalDevices/MedicalDevicesAdvisoryCommittee/CirculatorySystemDevicesPanel/ucm240575.htm>.
2. FDA will respond separately to the reclassification petition and will address the issues raised in that petition in its response; certain issues, however, may be addressed in both this document and the petition response due to the overlapping discussions in those documents. The reclassification petition is available at <http://www.regulations.gov/#!documentDetail;D=FDA-2013-N-0234-0002>.
3. U.S. Food and Drug Administration, Medical Device Recalls Database, available at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRES/res.cfm>.

4. Requests for Feedback on Medical Device Submissions: The Pre-Submission Program and Meetings with Food and Drug Administration Staff, Guidance for Industry and Food and Drug Administration Staff, February 18, 2014, available at <http://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM311176.pdf>.
5. Acceptance and Filing Reviews for Premarket Approval Applications (PMAs), Guidance for Industry and Food and Drug Administration Staff, December 31, 2012, available at <http://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM313368.pdf>.
6. Guidance for Industry and FDA Staff: Bundling Multiple Devices or Multiple Indications in a Single Submission, June 22, 2007, available at <http://www.fda.gov/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm089731.htm?>

List of Subjects in 21 CFR Part 870

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 870 is amended as follows:

PART 870—CARDIOVASCULAR DEVICES

- 1. The authority citation for 21 CFR part 870 continues to read as follows:
Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 371.
- 2. Section 870.5310 is amended by revising the section heading and paragraphs (a) and (c) to read as follows:

§ 870.5310 Automated external defibrillator system.

(a) *Identification.* An automated external defibrillator (AED) system consists of an AED and those accessories necessary for the AED to detect and interpret an electrocardiogram and deliver an electrical shock (e.g., battery, pad electrode, adapter, and hardware key for pediatric use). An AED system analyzes the patient's electrocardiogram, interprets the cardiac rhythm, and automatically delivers an electrical shock (fully automated AED), or advises the user to deliver the shock (semi-automated or shock advisory AED) to treat ventricular fibrillation or pulseless ventricular tachycardia.

* * * * *

(c) *Date PMA or notice of completion of PDP is required.* A PMA will be required to be submitted to the Food and Drug Administration by May 4, 2015, for any AED that was in commercial distribution before May 28, 1976, or that has, by May 4, 2015, been found to be substantially equivalent to any AED that was in commercial distribution before May 28, 1976. A PMA will be required to be submitted to the Food and Drug Administration by May 4, 2015, for any AED accessory described in paragraph (a) of this section that was in commercial distribution before May 28, 1976, or that has, by May 4, 2015, been found to be substantially equivalent to any AED accessory described in paragraph (a) that was in commercial distribution before May 28, 1976. Any other AED and AED accessory described in paragraph (a), shall have an approved PMA or declared completed PDP in effect before being placed in commercial distribution.

Dated: January 28, 2015.

Leslie Kux,

Associate Commissioner for Policy.

[FR Doc. 2015-02049 Filed 2-2-15; 8:45 am]

BILLING CODE 4164-01-P

POSTAL SERVICE

39 CFR Part 20

International Product and Price Changes

AGENCY: Postal Service™.

ACTION: Final rule.

SUMMARY: The Postal Service is revising *Mailing Standards of the United States Postal Service*, International Mail Manual (IMM®), to reflect the prices, product features, and classification

changes to Competitive Services, as established by the Governors of the Postal Service.

DATES: *Effective date:* April 26, 2015.

FOR FURTHER INFORMATION CONTACT: Paula Rabkin at 202-268-2537.

SUPPLEMENTARY INFORMATION: New prices will be posted under Docket Number CP2015-33 on the Postal Regulatory Commission's Web site at <http://www.prc.gov>.

This final rule describes the international price and classification changes and the corresponding mailing standards changes for the following Competitive Services:

- Global Express Guaranteed® (GXG®).
- Priority Mail Express International™.
- Priority Mail International®.
- First-Class Package International Service™.
- International Priority Airmail™ (IPA®).
- International Surface Air Lift® (ISAL®).
- Direct Sacks of Printed Matter to One Addressee (Airmail M-bags).
- International Extra Services:
 - Certificate of Mailing.
 - Registered Mail™ Service.
 - Return Receipt Service.

New prices will be located on the Postal Explorer® Web site at <http://pe.usps.com>.

Global Express Guaranteed

Global Express Guaranteed (GXG) is the Postal Service's premier international expedited product provided through an alliance with FedEx Express®. The price increase for GXG service averages 7.2 percent.

The Postal Service continues to provide Commercial Base pricing to online customers who prepare and pay for GXG shipments via USPS-approved payment methods, with variable discounts up to 16 percent off the published retail prices for GXG.

The Postal Service also continues to offer Commercial Plus pricing price incentives for large volume customers who commit to tendering \$100,000 in annual postal revenue from GXG, Priority Mail Express International (PMEI), Priority Mail International (PMI), and First-Class Package International Service (FCPIS®) via USPS-approved payment methods, with variable discounts up to 24 percent off the published retail prices for GXG.

Priority Mail Express International

Priority Mail Express International (PMEI) service provides fast service to approximately 180 countries. A money-

back guarantee service (exceptions apply) is available for certain destinations. The price increase for PMEI service averages 6.7 percent. The Commercial Base price and Commercial Plus price for customers that prepare and pay for PMEI shipments via permit imprint, online at USPS.com®, or as registered end-users using an authorized PC Postage vendor will remain a variable discount (based on the item's weight and price group) of up to 13 percent below the retail price for Commercial Base price and up to 25 percent below the retail price for Commercial Plus price.

The Postal Service continues to offer PMEI Commercial Plus pricing that includes discount price incentives to large volume customers who commit to tendering at least \$100,000 in annual postal revenue from GXG, PMEI, Priority Mail International and First-Class Package International Service. The Postal Service will continue to include PMEI in customized Global Expedited Package Services (GEPS) contracts offered to customers who meet certain revenue thresholds and are willing to commit to a larger amount of postal revenue for PMEI and Priority Mail International.

Priority Mail International

Priority Mail International (PMI) is a way to send merchandise and documents to about 180 countries. The price increase for PMI service averages 6.8 percent. The Commercial Base price and Commercial Plus price for customers that prepare and pay for PMI items via permit imprint, online at USPS.com, or as registered end-users using an authorized PC Postage vendor will remain a variable discount (based on the item's weight and price group) of up to 13 percent below the retail price for Commercial Base price and up to 21 percent below the retail price for Commercial Plus price. Large volume mailers who commit to tendering at least \$100,000 in annual postal revenue from GXG, PMEI, PMI, and First-Class Package International Service may request authorization for Commercial Plus discount prices. The Postal Service will continue to include PMI in customized Global Expedited Package Services (GEPS) contracts offered to customers who meet certain revenue thresholds and are willing to commit to a larger amount of revenue to the USPS® for PMEI and PMI.

In this filing we are proposing a structural change to create price zones for PMI to Canada. New zoned prices, based on the origin ZIP Code for PMI destined to Canada, will encourage customers to better use our network and